

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112423\
 Data File : VX039026.D
 Acq On : 24 Nov 2023 15:55
 Operator : JC/MD
 Sample : 04699-07 2.5PPB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Nov 27 03:48:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X112423W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Nov 27 03:22:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.897	128	9142	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	75404	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	92361	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	42055	33.626	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	= 112.100%#	
60) 4-Bromofluorobenzene	11.079	95	58955	29.224	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	= 97.400%	
63) Toluene-d8	8.646	98	112477	30.802	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	= 102.667%	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	3547	2.592	ug/l	91
3) Chloromethane	1.294	50	4301	2.745	ug/l	96
4) Vinyl Chloride	1.373	62	4530	2.864	ug/l	89
5) Bromomethane	1.605	94	3272	3.382	ug/l	92
6) Chloroethane	1.690	64	2801	3.041	ug/l	98
7) Trichlorofluoromethane	1.886	101	6107	2.797	ug/l	94
8) Diethyl Ether	2.135	74	2414	2.881	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.331	101	3439	2.788	ug/l	98
10) 1,1-Dichloroethene	2.318	96	3539	2.846	ug/l	91
11) Methyl Iodide	2.452	142	3585	2.631	ug/l	89
12) Methyl Acetate	2.702	43	11583	2.987	ug/l	99
13) Acrolein	2.233	56	5299	15.596	ug/l	95
14) Acrylonitrile	3.062	53	13478	14.209	ug/l	99
15) Acetone	2.379	58	4726	13.706	ug/l	94
16) Carbon Disulfide	2.513	76	10705	2.842	ug/l	98
17) Allyl chloride	2.666	41	7428	3.009	ug/l	96
18) Methylene Chloride	2.788	84	4925	3.341	ug/l	93
19) trans-1,2-Dichloroethene	3.093	96	3845	2.734	ug/l	90
20) Diisopropyl ether	3.757	45	13290	2.829	ug/l #	84
21) 1,1-Dichloroethane	3.605	63	7542	2.888	ug/l	99
22) cis-1,2-Dichloroethene	4.489	96	2491	1.528	ug/l #	8
23) tert-Butyl Alcohol	2.946	59	9688	18.449	ug/l #	100
24) Methyl tert-Butyl Ether	3.111	73	12798	2.794	ug/l	92
25) Chloroform	5.098	83	4581	1.747	ug/l	91
26) Cyclohexane	5.464	56	7195	3.044	ug/l #	91
29) 1,1-Dichloropropene	5.696	75	5815	2.751	ug/l	83
30) 2-Butanone	4.562	43	21532	13.530	ug/l	97
31) 2,2-Dichloropropane	4.476	77	5932	2.454	ug/l #	77
32) 1,1,1-Trichloroethane	5.391	97	6318	2.586	ug/l	94
33) Carbon Tetrachloride	5.671	117	4781	2.323	ug/l	91
34) Benzene	6.043	78	15735	2.555	ug/l #	88
35) Methacrylonitrile	4.922	41	3371	2.391	ug/l	92
36) 1,2-Dichloroethane	6.098	62	5948	2.633	ug/l	96
37) Trichloroethene	7.128	130	4045	2.588	ug/l	99
38) Methylcyclohexane	7.385	83	6609	2.509	ug/l	95
39) 1,2-Dichloropropane	7.433	63	4253	2.619	ug/l	99
40) Dibromomethane	7.580	93	2847	2.481	ug/l	98
41) Bromodichloromethane	7.823	83	5598	2.540	ug/l	93
42) Vinyl Acetate	3.727	43	48030	12.671	ug/l	99


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112423\
 Data File : VX039026.D
 Acq On : 24 Nov 2023 15:55
 Operator : JC/MD
 Sample : 04699-07 2.5PPB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Nov 27 03:48:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X112423W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Nov 27 03:22:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) Ethyl Acetate	4.720	43	6114	2.233	ug/l	98
44) Isopropyl Acetate	6.342	43	10625	2.462	ug/l	100
45) 1,4-Dioxane	7.665	88	2367	53.720	ug/l	95
46) Methyl methacrylate	7.695	41	6049	2.451	ug/l	98
47) n-amyl Acetate	10.841	43	7700	2.108	ug/l	94
48) t-1,3-Dichloropropene	8.982	75	6038	2.362	ug/l	91
49) cis-1,3-Dichloropropene	8.372	75	6607	2.506	ug/l #	78
50) 1,1,2-Trichloroethane	9.152	97	3931	2.522	ug/l	91
51) Ethyl methacrylate	9.122	69	4746	2.432	ug/l	93
52) 1,3-Dichloropropane	9.311	76	6850	2.563	ug/l	100
53) Dibromochloromethane	9.518	129	3790	2.264	ug/l	99
54) 1,2-Dibromoethane	9.610	107	4139	2.421	ug/l	96
55) 2-Chloroethyl vinyl ether	8.244	63	15711	12.704	ug/l	99
56) Bromoform	10.799	173	2530	2.094	ug/l	100
58) 4-Methyl-2-Pentanone	8.573	43	38102	13.604	ug/l	98
59) 2-Hexanone	9.433	43	28936	12.586	ug/l	99
61) Tetrachloroethene	9.274	164	3164	2.635	ug/l	96
62) Toluene	8.720	91	17374	2.650	ug/l	99
64) Chlorobenzene	10.079	112	10670	2.675	ug/l	91
65) 1,1,1,2-Tetrachloroethane	10.158	131	3574	2.488	ug/l	99
66) Ethyl Benzene	10.195	91	18951	2.538	ug/l	100
67) m/p-Xylenes	10.299	106	14132	5.122	ug/l	98
68) o-Xylene	10.640	106	6862	2.567	ug/l	96
69) Styrene	10.658	104	9154	2.314	ug/l	93
70) Isopropylbenzene	10.963	105	18356	2.567	ug/l	99
71) 1,1,2,2-Tetrachloroethane	11.213	83	6715	2.515	ug/l	97
72) 1,2,3-Trichloropropane	11.237	75	7717	3.416	ug/l	90
73) Bromobenzene	11.201	156	4016	2.467	ug/l	99
74) n-propylbenzene	11.305	91	22685	2.589	ug/l	96
75) 2-Chlorotoluene	11.366	91	13175	2.563	ug/l	99
76) 1,3,5-Trimethylbenzene	11.451	105	14653	2.379	ug/l	97
77) t-1,4-Dichloro-2-butene	11.018	75	1664	1.859	ug/l	81
78) 4-Chlorotoluene	11.457	91	15270	2.527	ug/l	99
79) tert-butylbenzene	11.713	119	13970	2.385	ug/l	93
80) 1,2,4-Trimethylbenzene	11.756	105	14418	2.338	ug/l	100
81) sec-Butylbenzene	11.890	105	18115	2.446	ug/l	100
82) p-Isopropyltoluene	12.012	119	14762	2.352	ug/l	99
83) 1,3-Dichlorobenzene	11.969	146	7555	2.425	ug/l	97
84) 1,4-Dichlorobenzene	12.042	146	8183	2.560	ug/l	97
85) n-Butylbenzene	12.335	91	15126	2.440	ug/l	93
86) Hexachloroethane	12.536	117	2535	2.132	ug/l	91
87) 1,2-Dichlorobenzene	12.335	146	7379	2.463	ug/l	96
88) 1,2-Dibromo-3-Chloropr...	12.944	75	1656	2.426	ug/l	90
89) 1,2,4-Trichlorobenzene	13.591	180	4698	2.365	ug/l	96
90) Hexachlorobutadiene	13.719	225	1567	2.209	ug/l	97
91) Naphthalene	13.780	128	17472	2.385	ug/l	98
92) 1,2,3-Trichlorobenzene	13.963	180	4546	2.363	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Bromochloromethane

Concen: 30.000 ug/l

RT: 4.897 min Scan# 61

Delta R.T. 0.000 min


Lab File: VX039026.D

Acq: 24 Nov 2023 15:55

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion:128 Resp: 9142

Ion Ratio Lower Upper

128 100

49 188.5 0.0 444.0

130 123.3 0.0 297.8



#2

Dichlorodifluoromethane

Concen: 2.592 ug/l

RT: 1.166 min Scan# 13

Delta R.T. -0.000 min

Lab File: VX039026.D




Acq: 24 Nov 2023 15:55


Tgt Ion: 85 Resp: 3547

Ion Ratio Lower Upper

85 100

87 38.7 16.7 50.1





#3

Chloromethane

Concen: 2.745 ug/l

RT: 1.294 min Scan# 34

Delta R.T. -0.000 min

Lab File: VX039026.D

Acq: 24 Nov 2023 15:55

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 50 Resp: 4301

Ion Ratio Lower Upper

50 100

52 37.0 27.8 41.8



#4

Vinyl Chloride

Concen: 2.864 ug/l

RT: 1.373 min Scan# 47

Delta R.T. -0.000 min

Lab File: VX039026.D




Acq: 24 Nov 2023 15:55


Tgt Ion: 62 Resp: 4530

Ion Ratio Lower Upper

62 100

64 26.5 26.3 39.5





#5

Bromomethane

Concen: 3.382 ug/l

RT: 1.605 min Scan# 81

Delta R.T. -0.006 min

Lab File: VX039026.D

Acq: 24 Nov 2023 15:55

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 94 Resp: 3272

Ion Ratio Lower Upper

94 100

96 101.7 74.9 112.3



#6

Chloroethane

Concen: 3.041 ug/l

RT: 1.690 min Scan# 99

Delta R.T. -0.000 min

Lab File: VX039026.D




Acq: 24 Nov 2023 15:55


Tgt Ion: 64 Resp: 2801

Ion Ratio Lower Upper

64 100

66 30.1 25.0 37.6









#7
 Trichlorofluoromethane
 Concen: 2.797 ug/l
 RT: 1.886 min Scan# 11
 Delta R.T. -0.006 min
 Lab File: VX039026.D
 Acq: 24 Nov 2023 15:55




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion:101 Resp: 6107
 Ion Ratio Lower Upper
 101 100
 103 65.8 49.1 73.7



#8
 Diethyl Ether
 Concen: 2.881 ug/l
 RT: 2.135 min Scan# 172
 Delta R.T. 0.006 min
 Lab File: VX039026.D
 Acq: 24 Nov 2023 15:55

Tgt Ion: 74 Resp: 2414
 Ion Ratio Lower Upper
 74 100
 45 103.9 20.5 184.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 2.788 ug/l
RT: 2.331 min Scan# 204
Delta R.T. -0.000 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:101 Resp: 3439
Ion Ratio Lower Upper
101 100
85 47.8 0.0 95.2
151 65.7 0.0 136.8



#10
1,1-Dichloroethene
Concen: 2.846 ug/l
RT: 2.318 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55


Tgt Ion: 96 Resp: 3539
Ion Ratio Lower Upper
96 100
61 160.5 138.3 207.5
98 69.5 49.8 74.6








#11
Methyl Iodide
Concen: 2.631 ug/l
RT: 2.452 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:142 Resp: 3585
Ion Ratio Lower Upper
142 100
127 39.3 23.8 71.4
141 16.5 6.9 20.6




#12
Methyl Acetate
Concen: 2.987 ug/l
RT: 2.702 min Scan# 265
Delta R.T. 0.006 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55




Tgt Ion: 43 Resp: 11583
Ion Ratio Lower Upper
43 100
74 21.6 17.8 26.6








#13
Acrolein
Concen: 15.596 ug/l
RT: 2.233 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 56 Resp: 5299
Ion Ratio Lower Upper
56 100
55 75.4 57.2 85.8



#14
Acrylonitrile
Concen: 14.209 ug/l
RT: 3.062 min Scan# 324
Delta R.T. 0.006 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55

Tgt Ion: 53 Resp: 13478
Ion Ratio Lower Upper
53 100
52 82.6 41.9 125.6
51 37.8 18.3 54.9





#15

Acetone

Concen: 13.706 ug/l

RT: 2.379 min Scan# 211

Delta R.T. 0.006 min

Lab File: VX039026.D

Acq: 24 Nov 2023 15:55

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 58 Resp: 4726

Ion Ratio Lower Upper

58 100

43 301.1 249.9 374.9



#16

Carbon Disulfide

Concen: 2.842 ug/l

RT: 2.513 min Scan# 234

Delta R.T. -0.000 min

Lab File: VX039026.D




Acq: 24 Nov 2023 15:55


Tgt Ion: 76 Resp: 10705

Ion Ratio Lower Upper

76 100

78 10.8 8.0 12.0









#17
 Allyl chloride
 Concen: 3.009 ug/l
 RT: 2.666 min Scan# 259
 Delta R.T. -0.000 min
 Lab File: VX039026.D
 Acq: 24 Nov 2023 15:55




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion:	41	Resp:	7428
Ion	Ratio	Lower	Upper
41	100		
39	78.0	36.5	109.5
76	39.4	19.7	59.0



#18
 Methylene Chloride
 Concen: 3.341 ug/l
 RT: 2.788 min Scan# 279
 Delta R.T. -0.000 min
 Lab File: VX039026.D
 Acq: 24 Nov 2023 15:55

Tgt Ion:	84	Resp:	4925
Ion	Ratio	Lower	Upper
84	100		
49	120.6	103.8	155.8
51	37.2	31.8	47.6
86	58.7	50.8	76.2









#19
trans-1,2-Dichloroethene
Concen: 2.734 ug/l
RT: 3.093 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 96 Resp: 3845
Ion Ratio Lower Upper
96 100
61 135.9 117.9 176.9
98 53.4 51.0 76.6



#20
Diisopropyl ether
Concen: 2.829 ug/l
RT: 3.757 min Scan# 438
Delta R.T. -0.000 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55

Tgt Ion: 45 Resp: 13290
Ion Ratio Lower Upper
45 100
43 104.8 67.7 101.5#
87 27.6 21.5 32.3
59 11.4 10.9 16.3





#21

1,1-Dichloroethane

Concen: 2.888 ug/l

RT: 3.605 min Scan# 41

Delta R.T. -0.006 min

Lab File: VX039026.D





Acq: 24 Nov 2023 15:55

Instrument :

MSVOA_X

ClientSampleId :

Tgt Ion:	63	Resp:	7542
Ion	Ratio	Lower	Upper
63	100		
98	6.1	2.9	8.8
100	3.0	2.0	5.8



#22

cis-1,2-Dichloroethene

Concen: 1.528 ug/l




RT: 4.489 min Scan# 558


Delta R.T. -0.006 min

Lab File: VX039026.D

Acq: 24 Nov 2023 15:55

Tgt Ion:	96	Resp:	2491
Ion	Ratio	Lower	Upper
96	100		
61	284.0	119.8	179.6#
98	110.1	50.9	76.3#









#23
tert-Butyl Alcohol
Concen: 18.449 ug/l
RT: 2.946 min Scan# 305
Delta R.T. -0.000 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 59 Resp: 9688
Ion Ratio Lower Upper
59 100
52 0.0 0.0 0.0



#24
Methyl tert-Butyl Ether
Concen: 2.794 ug/l
RT: 3.111 min Scan# 332
Delta R.T. -0.000 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55

Tgt Ion: 73 Resp: 12798
Ion Ratio Lower Upper
73 100
43 25.1 17.0 25.6









#25
Chloroform
Concen: 1.747 ug/l
RT: 5.098 min Scan# 61
Delta R.T. 0.006 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 4581
Ion Ratio Lower Upper
83 100
85 60.9 54.6 81.8



#26
Cyclohexane
Concen: 3.044 ug/l
RT: 5.464 min Scan# 718
Delta R.T. -0.006 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55

Tgt Ion: 56 Resp: 7195
Ion Ratio Lower Upper
56 100
89 0.0 0.0 0.0
84 75.7 67.2 100.8





#27

1,2-Dichloroethane-d4

Concen: 33.626 ug/l

RT: 5.958 min Scan# 798

Delta R.T. 0.006 min

Lab File: VX039026.D

Acq: 24 Nov 2023 15:55

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 65 Resp: 42055

Ion Ratio Lower Upper

65 100

67 52.9 42.5 63.7



#28

1,4-Difluorobenzene

Concen: 30.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. -0.000 min

Lab File: VX039026.D

Acq: 24 Nov 2023 15:55




Tgt Ion: 114 Resp: 75404


Ion Ratio Lower Upper

114 100

63 21.9 17.0 25.4

88 17.2 12.7 19.1





#29

1,1-Dichloropropene

Concen: 2.751 ug/l

RT: 5.696 min Scan# 755

Delta R.T. 0.006 min

Lab File: VX039026.D

Acq: 24 Nov 2023 15:55

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 75 Resp: 5815

Ion Ratio Lower Upper

75 100

110 25.2 0.0 71.2

77 22.2 0.0 62.6



#30

2-Butanone

Concen: 13.530 ug/l

RT: 4.562 min Scan# 570

Delta R.T. 0.012 min

Lab File: VX039026.D

Acq: 24 Nov 2023 15:55




Tgt Ion: 43 Resp: 21532


Ion Ratio Lower Upper

43 100

57 8.9 6.6 10.0

72 23.9 20.6 31.0





#31

2,2-Dichloropropane

Concen: 2.454 ug/l

RT: 4.476 min Scan# 51

Delta R.T. -0.000 min

Lab File: VX039026.D

Acq: 24 Nov 2023 15:55

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 77 Resp: 5932

Ion Ratio Lower Upper

77 100

97 11.7 18.2 27.4#



#32

1,1,1-Trichloroethane

Concen: 2.586 ug/l

RT: 5.391 min Scan# 706

Delta R.T. 0.012 min

Lab File: VX039026.D

Acq: 24 Nov 2023 15:55




Tgt Ion: 97 Resp: 6318


Ion Ratio Lower Upper

97 100

99 60.4 52.6 79.0


61 49.7 36.9 55.3








#33
Carbon Tetrachloride
Concen: 2.323 ug/l
RT: 5.671 min Scan# 711
Delta R.T. -0.006 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:117 Resp: 4781
Ion Ratio Lower Upper
117 100
119 87.5 78.2 117.4
121 33.9 24.7 37.1



#34
Benzene
Concen: 2.555 ug/l
RT: 6.043 min Scan# 813
Delta R.T. 0.006 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55


Tgt Ion: 78 Resp: 15735
Ion Ratio Lower Upper
78 100
77 17.5 19.3 28.9#
51 13.3 14.4 21.6#








#35
Methacrylonitrile
Concen: 2.391 ug/l
RT: 4.922 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 41	Resp: 3371
Ion Ratio	Lower Upper
41 100	
39 50.1	28.2 84.7
67 60.7	34.5 103.5
52 31.4	16.3 48.8



#36
1,2-Dichloroethane
Concen: 2.633 ug/l
RT: 6.098 min Scan# 822
Delta R.T. 0.018 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55

Tgt Ion: 62	Resp: 5948
Ion Ratio	Lower Upper
62 100	
98 7.2	7.0 10.4





#37

Trichloroethene

Concen: 2.588 ug/l

RT: 7.128 min Scan# 990

Delta R.T. 0.006 min

Lab File: VX039026.D

Acq: 24 Nov 2023 15:55

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion:130 Resp: 4045

Ion Ratio Lower Upper

130 100

95 106.0 85.8 128.8



#38

Methylcyclohexane

Concen: 2.509 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. 0.006 min

Lab File: VX039026.D

Acq: 24 Nov 2023 15:55




Tgt Ion: 83 Resp: 6609


Ion Ratio Lower Upper

83 100

55 84.2 38.6 116.0

98 44.4 22.3 66.8









#39
1,2-Dichloropropane
Concen: 2.619 ug/l
RT: 7.433 min Scan# 1041
Delta R.T. 0.006 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 63 Resp: 4253
Ion Ratio Lower Upper
63 100
65 30.2 24.6 37.0



#40
Dibromomethane
Concen: 2.481 ug/l
RT: 7.580 min Scan# 1065
Delta R.T. -0.000 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55

Tgt Ion: 93 Resp: 2847
Ion Ratio Lower Upper
93 100
95 85.5 0.0 163.4
174 83.2 0.0 166.4









#41
Bromodichloromethane
Concen: 2.540 ug/l
RT: 7.823 min Scan# 1105
Delta R.T. 0.006 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 5598
Ion Ratio Lower Upper
83 100
85 57.5 50.8 76.2
127 8.6 5.9 8.9



#42
Vinyl Acetate
Concen: 12.671 ug/l
RT: 3.727 min Scan# 433
Delta R.T. 0.006 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55

Tgt Ion: 43 Resp: 48030
Ion Ratio Lower Upper
43 100
86 7.9 6.7 10.1









#43
Ethyl Acetate
Concen: 2.233 ug/l
RT: 4.720 min Scan# 596
Delta R.T. 0.006 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 6114
Ion Ratio Lower Upper
43 100
45 15.7 11.8 17.8



#44
Isopropyl Acetate
Concen: 2.462 ug/l
RT: 6.342 min Scan# 862
Delta R.T. 0.006 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55

Tgt Ion: 43 Resp: 10625
Ion Ratio Lower Upper
43 100
61 19.6 15.6 23.4









#45
1,4-Dioxane
Concen: 53.720 ug/l
RT: 7.665 min Scan# 1077
Delta R.T. 0.012 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 88	Resp: 2367
Ion Ratio	Lower Upper
88 100	
43 31.4	28.1 42.1
58 68.9	57.7 86.5



#46
Methyl methacrylate
Concen: 2.451 ug/l
RT: 7.695 min Scan# 1084
Delta R.T. 0.006 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55

Tgt Ion: 41	Resp: 6049
Ion Ratio	Lower Upper
41 100	
69 80.6	39.6 118.8
39 62.6	30.1 90.3









#47
n-amy1 Acetate
Concen: 2.108 ug/l
RT: 10.841 min Scan# 1100
Delta R.T. -0.000 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 7700
Ion Ratio Lower Upper
43 100
55 25.6 18.2 27.2



#48
t-1,3-Dichloropropene
Concen: 2.362 ug/l
RT: 8.982 min Scan# 1295
Delta R.T. 0.006 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55

Tgt Ion: 75 Resp: 6038
Ion Ratio Lower Upper
75 100
77 28.3 26.8 40.2





#49

cis-1,3-Dichloropropene

Concen: 2.506 ug/l

RT: 8.372 min Scan# 1195

Delta R.T. 0.006 min

Lab File: VX039026.D

Acq: 24 Nov 2023 15:55

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 75 Resp: 6607

Ion Ratio Lower Upper

75 100

77 36.6 24.4 36.6

39 70.5 40.6 60.8#



#50

1,1,2-Trichloroethane

Concen: 2.522 ug/l

RT: 9.152 min Scan# 1323

Delta R.T. -0.000 min

Lab File: VX039026.D

Acq: 24 Nov 2023 15:55

Tgt Ion: 97 Resp: 3931




Ion Ratio Lower Upper


97 100

83 96.5 69.0 103.6

85 62.7 46.3 69.5

99 69.0 50.9 76.3









#51
Ethyl methacrylate
Concen: 2.432 ug/l
RT: 9.122 min Scan# 11
Delta R.T. 0.006 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 69 Resp: 4746
Ion Ratio Lower Upper
69 100
41 62.6 36.2 108.6
39 42.4 21.1 63.4



#52
1,3-Dichloropropane
Concen: 2.563 ug/l
RT: 9.311 min Scan# 1349
Delta R.T. 0.006 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55

Tgt Ion: 76 Resp: 6850
Ion Ratio Lower Upper
76 100
78 32.5 0.0 65.0









#53
Dibromochloromethane
Concen: 2.264 ug/l
RT: 9.518 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:129 Resp: 3790
Ion Ratio Lower Upper
129 100
127 77.8 38.3 114.9



#54
1,2-Dibromoethane
Concen: 2.421 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. -0.000 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55

Tgt Ion:107 Resp: 4139
Ion Ratio Lower Upper
107 100
109 89.9 75.1 112.7





#55

2-Chloroethyl vinyl ether

Concen: 12.704 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. 0.006 min

Lab File: VX039026.D

Acq: 24 Nov 2023 15:55

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 63 Resp: 15711

Ion Ratio Lower Upper

63 100

65 31.9 25.8 38.6

106 25.0 20.6 31.0



#56

Bromoform

Concen: 2.094 ug/l

RT: 10.799 min Scan# 1593

Delta R.T. -0.000 min

Lab File: VX039026.D

Acq: 24 Nov 2023 15:55




Tgt Ion:173 Resp: 2530


Ion Ratio Lower Upper

173 100

175 46.9 37.6 56.4

254 6.8 5.8 8.8









#57
Chlorobenzene-d5
Concen: 30.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:117 Resp: 92361
Ion Ratio Lower Upper
117 100
82 61.0 47.0 70.6
119 32.5 25.2 37.8



#58
4-Methyl-2-Pentanone
Concen: 13.604 ug/l
RT: 8.573 min Scan# 1228
Delta R.T. 0.006 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55

Tgt Ion: 43 Resp: 38102
Ion Ratio Lower Upper
43 100
58 36.4 30.1 45.1
85 17.1 13.9 20.9





#59

2-Hexanone

Concen: 12.586 ug/l

RT: 9.433 min Scan# 1

Delta R.T. 0.006 min


Lab File: VX039026.D

Acq: 24 Nov 2023 15:55

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion: 43 Resp: 28936

Ion Ratio Lower Upper

43 100

58 52.8 42.7 64.1

57 17.6 14.8 22.2



#60

4-Bromofluorobenzene

Concen: 29.224 ug/l

RT: 11.079 min Scan# 1639

Delta R.T. -0.000 min

Lab File: VX039026.D

Acq: 24 Nov 2023 15:55




Tgt Ion: 95 Resp: 58955


Ion Ratio Lower Upper

95 100

174 63.8 55.4 83.2


176 64.7 52.6 78.8








#61
Tetrachloroethene
Concen: 2.635 ug/l
RT: 9.274 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:	164	Resp:	3164
Ion	Ratio	Lower	Upper
164	100		
166	137.1	104.0	156.0
129	100.6	82.0	123.0
131	97.5	76.4	114.6



#62
Toluene
Concen: 2.650 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55

Tgt Ion:	91	Resp:	17374
Ion	Ratio	Lower	Upper
91	100		
92	58.0	45.9	68.9









#63
Toluene-d8
Concen: 30.802 ug/l
RT: 8.646 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 98 Resp: 112477
Ion Ratio Lower Upper
98 100
100 64.0 51.0 76.4



#64
Chlorobenzene
Concen: 2.675 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55

Tgt Ion: 112 Resp: 10670
Ion Ratio Lower Upper
112 100
114 27.3 25.9 38.9









#65
1,1,1,2-Tetrachloroethane
Concen: 2.488 ug/l
RT: 10.158 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:131 Resp: 3574
Ion Ratio Lower Upper
131 100
133 94.5 0.0 191.8
119 68.9 0.0 135.4



#66
Ethyl Benzene
Concen: 2.538 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.006 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55

Tgt Ion: 91 Resp: 18951
Ion Ratio Lower Upper
91 100
106 30.4 24.4 36.6









#67
m/p-Xylenes
Concen: 5.122 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:106 Resp: 14132
Ion Ratio Lower Upper
106 100
91 209.1 169.4 254.0



#68
o-Xylene
Concen: 2.567 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55

Tgt Ion:106 Resp: 6862
Ion Ratio Lower Upper
106 100
91 229.4 111.2 333.6





#69

Styrene

Concen: 2.314 ug/l

RT: 10.658 min Scan# 11

Delta R.T. 0.006 min

Lab File: VX039026.D





Acq: 24 Nov 2023 15:55

Instrument :

MSVOA_X

ClientSampleId :

Tgt Ion:104	Resp:	9154
Ion Ratio	Lower	Upper
104	100	
78	61.1	43.7 65.5
103	59.9	44.6 67.0



#70

Isopropylbenzene

Concen: 2.567 ug/l




RT: 10.963 min Scan# 1620


Delta R.T. -0.000 min

Lab File: VX039026.D

Acq: 24 Nov 2023 15:55

Tgt Ion:105	Resp:	18356
Ion Ratio	Lower	Upper
105	100	
120	26.7	18.5 34.3









#71
 1,1,2,2-Tetrachloroethane
 Concen: 2.515 ug/l
 RT: 11.213 min Scan# 1661
 Delta R.T. -0.000 min
 Lab File: VX039026.D
 Acq: 24 Nov 2023 15:55




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion	Ratio	Lower	Upper
83	100		
131	9.5	4.6	13.8
85	62.3	32.5	97.4



#72
 1,2,3-Trichloropropane
 Concen: 3.416 ug/l
 RT: 11.237 min Scan# 1665
 Delta R.T. -0.000 min
 Lab File: VX039026.D
 Acq: 24 Nov 2023 15:55

Tgt Ion	Ratio	Lower	Upper
75	100		
77	32.4	0.0	76.4









#73
Bromobenzene
Concen: 2.467 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.006 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion	Ratio	Lower	Upper
156	100		
77	182.5	0.0	363.2
158	97.3	0.0	197.8



#74
n-propylbenzene
Concen: 2.589 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55

Tgt Ion	Ratio	Lower	Upper
91	100		
120	19.7	0.0	43.4









#75
2-Chlorotoluene
Concen: 2.563 ug/l
RT: 11.366 min Scan# 11
Delta R.T. 0.006 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 13175
Ion Ratio Lower Upper
91 100
126 31.6 0.0 64.2



#76
1,3,5-Trimethylbenzene
Concen: 2.379 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55

Tgt Ion: 105 Resp: 14653
Ion Ratio Lower Upper
105 100
120 48.4 0.0 92.6









#77
t-1,4-Dichloro-2-butene
Concen: 1.859 ug/l
RT: 11.018 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 75 Resp: 1664
Ion Ratio Lower Upper
75 100
53 112.3 46.9 140.8
89 55.0 21.3 63.7



#78
4-Chlorotoluene
Concen: 2.527 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.006 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55

Tgt Ion: 91 Resp: 15270
Ion Ratio Lower Upper
91 100
126 28.1 0.0 57.4









#79
tert-butylbenzene
Concen: 2.385 ug/l
RT: 11.713 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:119 Resp: 13970
Ion Ratio Lower Upper
119 100
91 69.2 0.0 125.6
134 24.0 0.0 43.8



#80
1,2,4-Trimethylbenzene
Concen: 2.338 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. 0.006 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55


Tgt Ion:105 Resp: 14418
Ion Ratio Lower Upper
105 100
120 43.3 0.0 86.4








#81
sec-Butylbenzene
Concen: 2.446 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. -0.000 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:105 Resp: 18115
Ion Ratio Lower Upper
105 100
134 19.1 0.0 38.4



#82
p-Isopropyltoluene
Concen: 2.352 ug/l
RT: 12.012 min Scan# 1792
Delta R.T. 0.006 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55

Tgt Ion:119 Resp: 14762
Ion Ratio Lower Upper
119 100
134 26.8 0.0 52.6
91 27.4 0.0 54.2





#83

1,3-Dichlorobenzene

Concen: 2.425 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min


Lab File: VX039026.D

Acq: 24 Nov 2023 15:55

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion:146 Resp: 7555

Ion Ratio Lower Upper

146 100

111 46.9 22.3 66.9

148 67.7 32.8 98.3



#84

1,4-Dichlorobenzene

Concen: 2.560 ug/l

RT: 12.042 min Scan# 1797

Delta R.T. -0.000 min

Lab File: VX039026.D

Acq: 24 Nov 2023 15:55




Tgt Ion:146 Resp: 8183


Ion Ratio Lower Upper

146 100

111 44.0 21.4 64.2

148 61.2 31.8 95.4





#85

n-Butylbenzene

Concen: 2.440 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.006 min

Lab File: VX039026.D

Acq: 24 Nov 2023 15:55

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 91 Resp: 15126

Ion Ratio Lower Upper

91 100

92 48.6 0.0 106.2

134 20.6 0.0 48.8



#86

Hexachloroethane

Concen: 2.132 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX039026.D




Acq: 24 Nov 2023 15:55


Tgt Ion: 117 Resp: 2535

Ion Ratio Lower Upper

117 100

201 56.8 31.8 95.3









#87
 1,2-Dichlorobenzene
 Concen: 2.463 ug/l
 RT: 12.335 min Scan# 1845
 Delta R.T. -0.000 min
 Lab File: VX039026.D
 Acq: 24 Nov 2023 15:55




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion:146 Resp: 7379
 Ion Ratio Lower Upper
 146 100
 111 49.1 22.4 67.2
 148 65.5 31.9 95.7



#88
 1,2-Dibromo-3-Chloropropane
 Concen: 2.426 ug/l
 RT: 12.944 min Scan# 1945
 Delta R.T. 0.006 min
 Lab File: VX039026.D
 Acq: 24 Nov 2023 15:55

Tgt Ion: 75 Resp: 1656
 Ion Ratio Lower Upper
 75 100
 155 63.3 58.2 87.4
 157 86.2 76.5 114.7





#89

1,2,4-Trichlorobenzene

Concen: 2.365 ug/l

RT: 13.591 min Scan# 2050

Delta R.T. 0.006 min

Lab File: VX039026.D

Acq: 24 Nov 2023 15:55

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:180 Resp: 4698

Ion Ratio Lower Upper

180 100

182 88.8 74.5 111.7

145 33.9 26.6 40.0



#90

Hexachlorobutadiene

Concen: 2.209 ug/l

RT: 13.719 min Scan# 2072

Delta R.T. -0.006 min

Lab File: VX039026.D

Acq: 24 Nov 2023 15:55




Tgt Ion:225 Resp: 1567


Ion Ratio Lower Upper

225 100

223 66.4 0.0 124.4

227 61.4 0.0 124.4









#91
Naphthalene
Concen: 2.385 ug/l
RT: 13.780 min Scan# 2081
Delta R.T. 0.006 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55

Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion:128 Resp: 17472
Ion Ratio Lower Upper
128 100
127 11.1 10.2 15.2
129 10.6 8.6 12.8



#92
1,2,3-Trichlorobenzene
Concen: 2.363 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. 0.006 min
Lab File: VX039026.D
Acq: 24 Nov 2023 15:55

Tgt Ion:180 Resp: 4546
Ion Ratio Lower Upper
180 100
182 94.4 0.0 191.6
145 37.9 0.0 69.6

