

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112425\
 Data File : VX048666.D
 Acq On : 24 Nov 2025 16:09
 Operator : JC/MD
 Sample : Q3710-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 COMPRESSOR-EXCAVATION

Quant Time: Nov 26 00:27:01 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:08:11 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	193235	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	385610	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	412630	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	215359	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	155205	57.642	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	115.280%	
35) Dibromofluoromethane	5.373	113	129613	44.410	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	88.820%	
50) Toluene-d8	8.635	98	465698	51.162	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	102.320%	
62) 4-Bromofluorobenzene	11.067	95	192169	56.651	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	113.300%	
Target Compounds						
					Qvalue	
16) Acetone	2.374	43	11574	9.081	ug/l	99
20) Methylene Chloride	2.794	84	7233	2.760	ug/l	97
25) 2-Butanone	4.556	43	11321	6.004	ug/l	98
43) Isopropyl Acetate	6.336	43	11304	1.462	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112425\
Data File : VX048666.D
Acq On : 24 Nov 2025 16:09
Operator : JC/MD
Sample : Q3710-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
COMPRESSOR-EXCAVATION

Quant Time: Nov 26 00:27:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 05:08:11 2025
Response via : Initial Calibration

