

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112425\
 Data File : VX048669.D
 Acq On : 24 Nov 2025 17:10
 Operator : JC/MD
 Sample : Q3711-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 72-12017

Quant Time: Nov 26 00:28:01 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:08:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	167033	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	359372	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	398811	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	205142	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	133131	57.200	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	114.400%
35) Dibromofluoromethane	5.379	113	111706	41.069	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	82.140%
50) Toluene-d8	8.635	98	421317	49.666	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.340%
62) 4-Bromofluorobenzene	11.061	95	181824	57.514	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	115.020%
Target Compounds						
					Qvalue	
16) Acetone	2.373	43	12956	11.760	ug/l	97
20) Methylene Chloride	2.788	84	7975	3.521	ug/l	96
25) 2-Butanone	4.562	43	11084	6.801	ug/l	100
43) Isopropyl Acetate	6.330	43	12703	1.763	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112425\
Data File : VX048669.D
Acq On : 24 Nov 2025 17:10
Operator : JC/MD
Sample : Q3711-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
72-12017

Quant Time: Nov 26 00:28:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 05:08:11 2025
Response via : Initial Calibration

