


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX112519\
Data File : VX013575.D
Acq On : 25 Nov 2019 09:28
Operator : JC/SP
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
Title : SW846 8260
Last Update : Sat Nov 02 07:06:38 2019



AutoFind: Scans 1679, 1680, 1681; Background Corrected with Scan 1672

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.1	27248	PASS
75	95	30	60	49.5	70616	PASS
95	95	100	100	100.0	142781	PASS
96	95	5	9	6.7	9547	PASS
173	174	0.00	2	1.1	1341	PASS
174	95	50	100	84.3	120303	PASS
175	174	5	9	7.9	9471	PASS
176	174	95	101	95.3	114694	PASS
177	176	5	9	6.6	7547	PASS