

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112520\
 Data File : VX019664.D
 Acq On : 25 Nov 2020 18:14
 Operator : JC/MD
 Sample : PB133264TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB133264TB

Quant Time: Nov 26 05:00:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	292968	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.82	114	483326	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	443673	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	182988	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		6.03	65	181593	47.81	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.62%	
35) Dibromofluoromethane		5.46	113	143398	47.27	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.54%	
50) Toluene-d8		8.70	98	588483	49.92	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.84%	
62) 4-Bromofluorobenzene		11.12	95	203874	48.07	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	96.14%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Chloromethane	1.28	50	2066	0.632	ug/l #	41
5) Bromomethane	1.62	94	433	0.325	ug/l #	80
6) Chloroethane	1.68	64	2218	Below Cal	#	46
11) Tert butyl alcohol	2.99	59	5066	7.047	ug/l #	73
13) Acrolein	2.28	56	192	0.397	ug/l #	64
16) Acetone	2.42	43	11385	7.750	ug/l	92
18) Methyl Acetate	2.75	43	5003	2.006	ug/l	97
20) Methylene Chloride	2.83	84	9637	2.625	ug/l	97
25) 2-Butanone	4.64	43	2016	0.914	ug/l	93
29) Tetrahydrofuran	5.10	42	1489	1.058	ug/l #	75
31) Cyclohexane	5.62	56	5705	1.155	ug/l #	1
36) 1,1-Dichloropropene	5.62	75	22129	4.884	ug/l #	49
38) Carbon Tetrachloride	5.63	117	28564	6.513	ug/l #	16
43) Isopropyl Acetate	6.41	43	81270	11.365	ug/l	99
48) Methyl methacrylate	7.65	41	4114	1.228	ug/l #	9
49) 1,4-Dioxane	7.85	88	197	2.591	ug/l #	1
51) 4-Methyl-2-Pentanone	8.59	43	78888	18.919	ug/l #	47
54) cis-1,3-Dichloropropene	8.59	75	41844	7.670	ug/l #	58
56) Ethyl methacrylate	9.29	69	1264	0.258	ug/l #	1
57) 1,3-Dichloropropane	9.52	76	7580	1.305	ug/l #	43
59) 2-Hexanone	9.42	43	77002	24.349	ug/l #	20
70) Styrene	10.70	104	275	1.671	ug/l	97
74) N-ethyl acetate	10.80	43	3466	0.731	ug/l #	42
76) 1,2,3-Trichloropropane	11.12	75	99859	25.006	ug/l #	37
80) 1,3,5-Trimethylbenzene	11.49	105	2672	0.252	ug/l	94
81) trans-1,4-Dichloro-2-buten	11.12	75	99859	66.054	ug/l #	12
95) Naphthalene	13.82	128	471	2.211	ug/l #	88


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112520\
Data File : VX019664.D
Acq On : 25 Nov 2020 18:14
Operator : JC/MD
Sample : PB133264TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB133264TB

Quant Time: Nov 26 05:00:54 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.62 min Scan# 744

Delta R.T. -0.01 min

Lab File: VX019664.D

Acq: 25 Nov 2020 18:14

Instrument :

MSVOA_X

ClientSampleId :


PB133264TB

Tot Ion:168 Resp: 292968

Ion Ratio Lower Upper

168 100

99 55.6 45.8 68.6



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.62

120000

100000

80000

60000



40000

20000

0

Time-->

5.50 5.60 5.70 5.80



#3

Chloromethane

Concen: 0.632 ug/l

RT: 1.28 min Scan# 31

Delta R.T. -0.04 min

Lab File: VX019664.D


Acq: 25 Nov 2020 18:14

Tgt Ion: 50 Resp: 2066

Ion Ratio Lower Upper

50 100

52 0.0 27.0 40.6#



Abundance Ion 49.90 (49.60 to 50.60): VX0

Ion 51.90 (51.60 to 52.60): VX0

1.28

800

600


400


200

0

Time-->

1.20 1.25 1.30 1.35





#5

Bromomethane

Concen: 0.325 ug/l

RT: 1.62 min Scan# 88

Delta R.T. 0.01 min

Lab File: VX019664.D

Acq: 25 Nov 2020 18:14

Instrument :

MSVOA_X

ClientSampled :


PB133264TB

Tot Ion: 94 Resp: 433

Ion Ratio Lower Upper




94 100

96 75.4 75.6 113.4#



Abundance Ion 93.90 (93.60 to 94.60): VX0

Ion 95.90 (95.60 to 96.60): VX0



#6

Chloroethane

Concen: Below Cal

RT: 1.68 min Scan# 97

Delta R.T. -0.01 min

Lab File: VX019664.D


Acq: 25 Nov 2020 18:14

Tgt Ion: 64 Resp: 2218

Ion Ratio Lower Upper



64 100


66 3.1 27.1 40.7#



Abundance Ion 63.90 (63.60 to 64.60): VX0

Ion 65.90 (65.60 to 66.60): VX0









#11

Tert butyl alcohol
 Concen: 7.047 ug/l
 RT: 2.99 min Scan# 312
 Delta R.T. -0.05 min
 Lab File: VX019664.D
 Acq: 25 Nov 2020 18:14

Instrument :
 MSVOA_X
 ClientSampleId :
 PB133264TB




Tot Ion: 59 Resp: 5066
 Ion Ratio Lower Upper
 59 100
 57 0.0 7.9 11.9#




#13

Acrolein
 Concen: 0.397 ug/l
 RT: 2.28 min Scan# 195
 Delta R.T. -0.00 min
 Lab File: VX019664.D
 Acq: 25 Nov 2020 18:14

Tgt Ion: 56 Resp: 192
 Ion Ratio Lower Upper
 56 100
 55 40.1 55.5 83.3#





#16

Acetone

Concen: 7.750 ug/l

RT: 2.42 min Scan# 219

Delta R.T. -0.01 min

Lab File: VX019664.D

Acq: 25 Nov 2020 18:14

Instrument :

MSVOA_X

ClientSampleId :





PB133264TB

Tot Ion: 43 Resp: 11385

Ion Ratio Lower Upper

43 100

58 38.7 27.4 41.2



#18

Methyl Acetate

Concen: 2.006 ug/l

RT: 2.75 min Scan# 273

Delta R.T. 0.00 min

Lab File: VX019664.D




Acq: 25 Nov 2020 18:14


Tgt Ion: 43 Resp: 5003

Ion Ratio Lower Upper

43 100

74 29.6 22.2 33.4









#20
Methvlene Chloride
Concen: 2.625 ug/l
RT: 2.83 min Scan# 286
Delta R.T. -0.00 min
Lab File: VX019664.D
Acq: 25 Nov 2020 18:14




Instrument :
MSVOA_X
ClientSampleId :
PB133264TB


Tot Ion:	84	Resp:	9637
Ion	Ratio	Lower	Upper
84	100		
49	108.9	88.5	132.7
51	33.6	28.0	42.0
86	62.4	53.0	79.4



#25
2-Butanone
Concen: 0.914 ug/l
RT: 4.64 min Scan# 583
Delta R.T. -0.00 min
Lab File: VX019664.D
Acq: 25 Nov 2020 18:14

Tgt Ion:	43	Resp:	2016
Ion	Ratio	Lower	Upper
43	100		
72	26.3	24.0	36.0





#29

Tetrahydrofuran

Concen: 1.058 ug/l

RT: 5.10 min Scan# 659

Delta R.T. 0.01 min

Lab File: VX019664.D

Acq: 25 Nov 2020 18:14

Instrument :

MSVOA_X

ClientSampleId :

PB133264TB


Tot Ion: 42 Resp: 1489

Ion Ratio Lower Upper

42 100

72 19.9 42.6 63.8#




71 47.5 39.0 58.6



Abundance Ion 42.00 (41.70 to 42.70): VX0

Ion 72.00 (71.70 to 72.70): VX0

Ion 71.00 (70.70 to 71.70): VX0



#31

Cyclohexane

Concen: 1.155 ug/l

RT: 5.62 min Scan# 744

Delta R.T. 0.09 min

Lab File: VX019664.D

Acq: 25 Nov 2020 18:14


Tgt Ion: 56 Resp: 5705

Ion Ratio Lower Upper

56 100

69 185.5 25.5 38.3#



84 157.0 71.1 106.7#




Abundance Ion 56.00 (55.70 to 56.70): VX0

Ion 69.00 (68.70 to 69.70): VX0

Ion 84.00 (83.70 to 84.70): VX0





#33

1,2-Dichloroethane-d4

Concen: 47.806 ug/l

RT: 6.03 min Scan# 810

Delta R.T. -0.01 min





Lab File: VX019664.D

Acq: 25 Nov 2020 18:14

Instrument :
MSVOA_X
ClientSampleId :
PB133264TB

Tot Ion: 65 Resp: 181593

Ion	Ratio	Lower	Upper
65	100		
67	52.7	0.0	105.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.82 min Scan# 941




Delta R.T. -0.01 min


Lab File: VX019664.D

Acq: 25 Nov 2020 18:14

Tgt Ion: 114 Resp: 483326

Ion	Ratio	Lower	Upper
114	100		
63	20.1	0.0	38.8
88	15.8	0.0	31.6









#35
 Dibromofluoromethane
 Concen: 47.266 ug/l
 RT: 5.46 min Scan# 717
 Delta R.T. -0.01 min
 Lab File: VX019664.D
 Acq: 25 Nov 2020 18:14




Instrument :
 MSVOA_X
 ClientSampleId :
 PB133264TB


Tot Ion:113 Resp: 143398
 Ion Ratio Lower Upper
 113 100
 111 102.4 81.5 122.3
 192 20.1 16.2 24.2



#36
 1,1-Dichloropropene
 Concen: 4.884 ug/l
 RT: 5.62 min Scan# 744
 Delta R.T. -0.14 min
 Lab File: VX019664.D
 Acq: 25 Nov 2020 18:14

Tgt Ion: 75 Resp: 22129
 Ion Ratio Lower Upper
 75 100
 110 8.7 18.1 54.3#
 77 0.0 24.2 36.4#





#38

Carbon Tetrachloride

Concen: 6.513 ug/l

RT: 5.63 min Scan# 745

Delta R.T. -0.12 min

Lab File: VX019664.D

Acq: 25 Nov 2020 18:14

Instrument :

MSVOA_X

ClientSampled :

PB133264TB


Tot Ion:117 Resp: 28564

Ion Ratio Lower Upper

117 100

119 4.5 75.9 113.9#




121 0.0 24.0 36.0#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V



#43

Isopropyl Acetate

Concen: 11.365 ug/l

RT: 6.41 min Scan# 873

Delta R.T. -0.01 min

Lab File: VX019664.D

Acq: 25 Nov 2020 18:14


Tgt Ion: 43 Resp: 81270

Ion Ratio Lower Upper

43 100

61 22.2 18.2 27.2



87 15.3 12.6 18.8




Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 61.00 (60.70 to 61.70): VX0

Ion 87.00 (86.70 to 87.70): VX0









#48
Methvl methacrylate
Concen: 1.228 ug/l
RT: 7.65 min Scan# 1077
Delta R.T. -0.09 min
Lab File: VX019664.D
Acq: 25 Nov 2020 18:14




Instrument :
MSVOA_X
ClientSampleId :
PB133264TB


Tot Ion: 41 Resp: 4114
Ion Ratio Lower Upper
41 100
69 0.0 77.0 115.6#
39 0.0 45.6 68.4#



#49
1,4-Dioxane
Concen: 2.591 ug/l
RT: 7.85 min Scan# 1110
Delta R.T. 0.13 min
Lab File: VX019664.D
Acq: 25 Nov 2020 18:14

Tgt Ion: 88 Resp: 197
Ion Ratio Lower Upper
88 100
43 709447.2 23.8 35.6#
58 3342.6 54.2 81.2#





#50

Toluene-d8

Concen: 49.917 ug/l

RT: 8.70 min Scan# 1248

Delta R.T. -0.00 min

Lab File: VX019664.D

Acq: 25 Nov 2020 18:14

Instrument :

MSVOA_X

ClientSampleId :





PB133264TB

Tot Ion: 98 Resp: 588483

Ion Ratio Lower Upper

98 100

100 66.6 52.4 78.6



#51

4-Methyl-2-Pentanone

Concen: 18.919 ug/l

RT: 8.59 min Scan# 1231

Delta R.T. -0.03 min

Lab File: VX019664.D




Acq: 25 Nov 2020 18:14


Tgt Ion: 43 Resp: 78888

Ion Ratio Lower Upper

43 100

58 9.9 35.1 52.7#









#54

cis-1,3-Dichloropropene
 Concen: 7.670 ug/l
 RT: 8.59 min Scan# 1231
 Delta R.T. 0.18 min
 Lab File: VX019664.D
 Acq: 25 Nov 2020 18:14

Instrument :
 MSVOA_X
 ClientSampleId :
 PB133264TB




Tot Ion: 75 Resp: 41844
 Ion Ratio Lower Upper
 75 100
 77 0.7 25.8 38.6#
 39 23.5 34.6 51.8#




#56

Ethyl methacrylate
 Concen: 0.258 ug/l
 RT: 9.29 min Scan# 1345
 Delta R.T. 0.13 min
 Lab File: VX019664.D
 Acq: 25 Nov 2020 18:14

Tgt Ion: 69 Resp: 1264
 Ion Ratio Lower Upper
 69 100
 41 5319.9 46.2 69.2#
 39 2094.4 25.5 38.3#





#57

1,3-Dichloropropane

Concen: 1.305 ug/l

RT: 9.52 min Scan# 1384

Delta R.T. 0.17 min

Lab File: VX019664.D

Acq: 25 Nov 2020 18:14

Instrument :

MSVOA_X

ClientSampleId :





PB133264TB

Tot Ion: 76 Resp: 7580

Ion Ratio Lower Upper

76 100

78 0.0 25.7 38.5#



#59

2-Hexanone

Concen: 24.349 ug/l

RT: 9.42 min Scan# 1367

Delta R.T. -0.05 min

Lab File: VX019664.D




Acq: 25 Nov 2020 18:14


Tgt Ion: 43 Resp: 77002

Ion Ratio Lower Upper

43 100

58 0.0 30.6 91.6#





#62

4-Bromofluorobenzene

Concen: 48.065 ug/l

RT: 11.12 min Scan# 1646

Delta R.T. -0.00 min

Lab File: VX019664.D

Acq: 25 Nov 2020 18:14

Instrument :

MSVOA_X

ClientSampleId :

PB133264TB





Tot Ion: 95 Resp: 203874

Ion Ratio Lower Upper

95 100

174 76.9 0.0 156.2

176 74.8 0.0 151.4



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.10 min Scan# 1478

Delta R.T. -0.00 min

Lab File: VX019664.D

Acq: 25 Nov 2020 18:14




Tgt Ion: 117 Resp: 443673


Ion Ratio Lower Upper

117 100

82 54.1 43.0 64.4

119 31.9 25.6 38.4









#70
Stvrene
Concen: 1.671 ug/l
RT: 10.70 min Scan# 1576
Delta R.T. 0.00 min
Lab File: VX019664.D
Acq: 25 Nov 2020 18:14




Instrument :
MSVOA_X
ClientSampled :
PB133264TB


Tot Ion:104 Resp: 275
Ion Ratio Lower Upper
104 100
78 47.6 39.5 59.3
103 56.7 43.6 65.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.06 min Scan# 1800
Delta R.T. -0.00 min
Lab File: VX019664.D
Acq: 25 Nov 2020 18:14

Tgt Ion:152 Resp: 182988
Ion Ratio Lower Upper
152 100
115 60.1 41.8 125.4
150 157.3 0.0 348.2





#74

N-amvl acetate

Concen: 0.731 ug/l

RT: 10.80 min Scan# 1594

Delta R.T. -0.07 min

Lab File: VX019664.D

Acq: 25 Nov 2020 18:14

Instrument :

MSVOA_X

ClientSampleId :

PB133264TB

Tot Ion: 43 Resp: 3466


Ion Ratio Lower Upper

43 100

70 0.0 39.7 59.5#

55 8.2 21.8 32.6#

61 0.0 22.2 33.4#






Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 70.00 (69.70 to 70.70): VX0

Ion 55.00 (54.70 to 55.70): VX0

Ion 61.00 (60.70 to 61.70): VX0



#76

1,2,3-Trichloropropane

Concen: 25.006 ug/l

RT: 11.12 min Scan# 1646

Delta R.T. -0.16 min

Lab File: VX019664.D


Acq: 25 Nov 2020 18:14

Tgt Ion: 75 Resp: 99859

Ion Ratio Lower Upper



75 100


77 1.2 20.0 60.0#



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 77.00 (76.70 to 77.70): VX0





#80

1,3,5-Trimethylbenzene

Concen: 0.252 ug/l

RT: 11.49 min Scan# 1706

Delta R.T. -0.00 min

Lab File: VX019664.D

Acq: 25 Nov 2020 18:14

Instrument :

MSVOA_X

ClientSampleId :


PB133264TB

Tot Ion:105 Resp: 2672

Ion Ratio Lower Upper




105 100

120 44.1 24.1 72.3



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V



#81

trans-1,4-Dichloro-2-butene

Concen: 66.054 ug/l

RT: 11.12 min Scan# 1646

Delta R.T. 0.07 min

Lab File: VX019664.D

Acq: 25 Nov 2020 18:14


Tgt Ion: 75 Resp: 99859

Ion Ratio Lower Upper

75 100

53 0.1 73.9 110.9#



89 0.0 37.3 55.9#




Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 53.00 (52.70 to 53.70): VX0

Ion 88.90 (88.60 to 89.60): VX0





#95

Naphthalene

Concen: 2.211 ug/l

RT: 13.82 min Scan# 2088

Delta R.T. 0.00 min

Lab File: VX019664.D

Acq: 25 Nov 2020 18:14

Instrument :

MSVOA_X

ClientSampleId :

PB133264TB


Tot Ion:128 Resp: 471

Ion Ratio Lower Upper

128 100

127 4.9 10.3 15.5#

129 12.1 8.7 13.1



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V

