

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112520\
 Data File : VX019655.D
 Acq On : 25 Nov 2020 14:45
 Operator : JC/MD
 Sample : L4835-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP02-20201119

Quant Time: Nov 26 06:10:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	298812	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	506140	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	468127	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	212402	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		6.03	65	190843	49.26	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.52%	
35) Dibromofluoromethane		5.46	113	149953	47.20	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.40%	
50) Toluene-d8		8.70	98	608214	49.27	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.54%	
62) 4-Bromofluorobenzene		11.12	95	227887	51.31	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.62%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.42	43	2905	1.939	ug/l	98
27) cis-1,2-Dichloroethene	4.56	96	3598	0.932	ug/l	98
44) Trichloroethene	7.18	130	7104	1.860	ug/l	97
95) Naphthalene	13.82	128	1770	2.286	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112520\
Data File : VX019655.D
Acq On : 25 Nov 2020 14:45
Operator : JC/MD
Sample : L4835-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP02-20201119

Quant Time: Nov 26 06:10:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

