

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112520\
 Data File : VX019657.D
 Acq On : 25 Nov 2020 15:31
 Operator : JC/MD
 Sample : L4840-06DL 1000X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 30738DL

Quant Time: Nov 26 06:15:49 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	306769	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	508826	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	445608	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	186510	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	189830	47.73	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.46%
35) Dibromofluoromethane	5.46	113	149537	46.82	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.64%
50) Toluene-d8	8.70	98	603798	48.65	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.30%
62) 4-Bromofluorobenzene	11.12	95	206044	46.14	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.28%

Target Compounds					Qvalue
16) Acetone	2.42	43	612032	397.894 ug/l	100
20) Methylene Chloride	2.82	84	3161	0.822 ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112520\
Data File : VX019657.D
Acq On : 25 Nov 2020 15:31
Operator : JC/MD
Sample : L4840-06DL 1000X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
30738DL

Quant Time: Nov 26 06:15:49 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

