

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112520\
 Data File : VX019658.D
 Acq On : 25 Nov 2020 15:55
 Operator : JC/MD
 Sample : L4841-03 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 16881

Quant Time: Nov 26 06:35:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	298621	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	493770	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.09	117	436696	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	199457	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		6.03	65	186664	48.21	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.42%	
35) Dibromofluoromethane		5.46	113	146063	47.13	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.26%	
50) Toluene-d8		8.70	98	583790	48.47	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.94%	
62) 4-Bromofluorobenzene		11.12	95	207090	47.79	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.58%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	3.03	59	10582	14.441	ug/l #	89
16) Acetone	2.43	43	219555	146.632	ug/l	99
20) Methylene Chloride	2.83	84	1966	0.525	ug/l #	90
25) 2-Butanone	4.65	43	14985	6.662	ug/l	99
31) Cyclohexane	5.55	56	2746	0.545	ug/l #	95
39) Methylcyclohexane	7.44	83	7289	1.352	ug/l	94
52) Toluene	8.76	92	4514	0.519	ug/l	98
64) Tetrachloroethene	9.32	164	5937	1.730	ug/l	95
67) Ethyl Benzene	10.23	91	5373	0.342	ug/l	89
68) m/p-Xylenes	10.34	106	11718	2.000	ug/l	99
69) o-Xylene	10.68	106	4591	0.820	ug/l	99
78) n-propylbenzene	11.34	91	6259	0.395	ug/l	99
80) 1,3,5-Trimethylbenzene	11.49	105	10269	0.887	ug/l	98
84) 1,2,4-Trimethylbenzene	11.79	105	40606	3.467	ug/l	88
89) n-Butylbenzene	12.37	91	5278	0.483	ug/l #	44
95) Naphthalene	13.82	128	50578	5.466	ug/l #	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112520\
Data File : VX019658.D
Acq On : 25 Nov 2020 15:55
Operator : JC/MD
Sample : L4841-03 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
16881

Quant Time: Nov 26 06:35:06 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

