

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112520\
 Data File : VX019662.D
 Acq On : 25 Nov 2020 17:28
 Operator : JC/MD
 Sample : L4835-25
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB01-20201119

Manual Integrations
 APPROVED

MMDadoda
 11/30/2020 9:54:03 AM

Quant Time: Nov 26 06:45:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	300791	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	492557	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	417478	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	155571	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	185012	47.44	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	94.88%		
35) Dibromofluoromethane	5.46	113	146226	47.30	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	94.60%		
50) Toluene-d8	8.70	98	583298	48.55	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	97.10%		
62) 4-Bromofluorobenzene	11.12	95	184354	42.65	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	85.30%		
Target Compounds						
16) Acetone	2.42	43	47963	31.801	ug/l	99
25) 2-Butanone	4.65	43	2297m	1.014	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112520\
Data File : VX019662.D
Acq On : 25 Nov 2020 17:28
Operator : JC/MD
Sample : L4835-25
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB01-20201119

Manual Integrations
APPROVED

MMDadoda
11/30/2020 9:54:03 AM

Quant Time: Nov 26 06:45:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

