

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112520\
 Data File : VX019665.D
 Acq On : 25 Nov 2020 18:38
 Operator : JC/MD
 Sample : L4835-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ROOF-DRAIN-20201119

Quant Time: Nov 26 06:51:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	301512	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	503200	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	444181	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	172461	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		6.02	65	189925	48.58	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.16%	
35) Dibromofluoromethane		5.46	113	149417	47.31	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.62%	
50) Toluene-d8		8.70	98	599469	48.84	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.68%	
62) 4-Bromofluorobenzene		11.12	95	200404	45.38	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.76%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.42	43	2717	1.797	ug/l	94
18) Methyl Acetate	2.75	43	4594	1.858	ug/l	100
27) cis-1,2-Dichloroethene	4.56	96	3703	0.951	ug/l	89
44) Trichloroethene	7.18	130	7268	1.914	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112520\
Data File : VX019665.D
Acq On : 25 Nov 2020 18:38
Operator : JC/MD
Sample : L4835-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ROOF-DRAIN-20201119

Quant Time: Nov 26 06:51:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

