

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112522\
 Data File : VX033062.D
 Acq On : 25 Nov 2022 12:59
 Operator : JC/MD
 Sample : N5741-12DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-31DL

Quant Time: Nov 28 00:46:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	139753	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	224338	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	197685	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	91961	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	72907	53.615	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	107.220%
35) Dibromofluoromethane	5.391	113	57712	45.632	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.260%
50) Toluene-d8	8.647	98	194941	51.284	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.560%
62) 4-Bromofluorobenzene	11.079	95	71279	42.904	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.800%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.940	59	5747	26.638	ug/l	# 72
19) Methyl tert-butyl Ether	3.111	73	131117	31.000	ug/l	100
22) Diisopropyl ether	3.757	45	1951	0.455	ug/l	92
40) Benzene	6.043	78	12678	2.236	ug/l	99
67) Ethyl Benzene	10.195	91	7774	1.112	ug/l	95
68) m/p-Xylenes	10.305	106	995	0.363	ug/l	99
69) o-Xylene	10.646	106	876	0.331	ug/l	92
73) Isopropylbenzene	10.963	105	2341	0.356	ug/l	99
78) n-propylbenzene	11.305	91	2236	0.298	ug/l	98
84) 1,2,4-Trimethylbenzene	11.756	105	1277	0.232	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112522\
Data File : VX033062.D
Acq On : 25 Nov 2022 12:59
Operator : JC/MD
Sample : N5741-12DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-31DL

Quant Time: Nov 28 00:46:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

