

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112522\
 Data File : VX033064.D
 Acq On : 25 Nov 2022 13:46
 Operator : JC/MD
 Sample : N5747-01DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-28DL

Quant Time: Nov 28 00:46:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	137452	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	219492	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	193619	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	89242	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	72190	53.999	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.000%	
35) Dibromofluoromethane	5.391	113	57335	46.335	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 92.660%	
50) Toluene-d8	8.647	98	195106	52.512	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 105.020%	
62) 4-Bromofluorobenzene	11.079	95	70626	43.450	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 86.900%	
Target Compounds						Qvalue
11) Tert butyl alcohol	2.989	59	26103	123.017	ug/l	97
19) Methyl tert-butyl Ether	3.117	73	6705	1.612	ug/l #	90
22) Diisopropyl ether	3.770	45	1766	0.419	ug/l	87
40) Benzene	6.044	78	72091	12.995	ug/l	99
52) Toluene	8.720	92	2334	0.640	ug/l	94
67) Ethyl Benzene	10.195	91	12268	1.792	ug/l	99
68) m/p-Xylenes	10.305	106	6297	2.348	ug/l	91
69) o-Xylene	10.640	106	3566	1.374	ug/l	94
73) Isopropylbenzene	10.964	105	3860	0.605	ug/l	99
78) n-propylbenzene	11.305	91	4573	0.627	ug/l	95
80) 1,3,5-Trimethylbenzene	11.451	105	1346	0.251	ug/l	96
84) 1,2,4-Trimethylbenzene	11.750	105	5888	1.102	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112522\
Data File : VX033064.D
Acq On : 25 Nov 2022 13:46
Operator : JC/MD
Sample : N5747-01DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-28DL

Quant Time: Nov 28 00:46:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

