

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112522\
 Data File : VX033068.D
 Acq On : 25 Nov 2022 15:18
 Operator : JC/MD
 Sample : N5747-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB112122

Quant Time: Nov 28 00:48:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	136064	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	217069	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	190835	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	87070	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	71210	53.797	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	107.600%
35) Dibromofluoromethane	5.391	113	56208	45.931	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.860%
50) Toluene-d8	8.647	98	191917	52.218	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.440%
62) 4-Bromofluorobenzene	11.079	95	67783	42.166	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.340%
Target Compounds						
16) Acetone	2.404	43	1271	2.054	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112522\
Data File : VX033068.D
Acq On : 25 Nov 2022 15:18
Operator : JC/MD
Sample : N5747-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB112122

Quant Time: Nov 28 00:48:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

