

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112524\
 Data File : VX043978.D
 Acq On : 25 Nov 2024 11:55
 Operator : JC/MD
 Sample : VX1125MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1125MBL01

Quant Time: Nov 26 00:44:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	114393	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	218626	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	191407	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	81880	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	101006	51.480	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.960%
35) Dibromofluoromethane	5.379	113	74906	47.949	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.900%
50) Toluene-d8	8.647	98	267140	49.608	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.220%
62) 4-Bromofluorobenzene	11.079	95	87309	47.640	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	95.280%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112524\
Data File : VX043978.D
Acq On : 25 Nov 2024 11:55
Operator : JC/MD
Sample : VX1125MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1125MBL01

Quant Time: Nov 26 00:44:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

