

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112524\
 Data File : VX043979.D
 Acq On : 25 Nov 2024 12:18
 Operator : JC/MD
 Sample : VX1125WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1125WBL01

Quant Time: Nov 26 00:44:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	113649	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	223629	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	208120	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	90683	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	101387	52.013	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	104.020%	
35) Dibromofluoromethane	5.379	113	75383	47.175	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	94.360%	
50) Toluene-d8	8.647	98	284104	51.578	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	103.160%	
62) 4-Bromofluorobenzene	11.079	95	95049	50.703	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	101.400%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112524\
Data File : VX043979.D
Acq On : 25 Nov 2024 12:18
Operator : JC/MD
Sample : VX1125WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1125WBL01

Quant Time: Nov 26 00:44:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

