

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112524\
 Data File : VX043982.D
 Acq On : 25 Nov 2024 13:31
 Operator : JC/MD
 Sample : P4846-01DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-MW-EVAL-03D-20241113DL

Quant Time: Nov 26 00:47:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	107746	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.757	114	202746	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	176451	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	73448	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	93012	50.330	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	100.660%	
35) Dibromofluoromethane	5.379	113	69517	47.985	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	95.980%	
50) Toluene-d8	8.647	98	247863	49.633	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	99.260%	
62) 4-Bromofluorobenzene	11.079	95	81837	48.152	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	96.300%	
Target Compounds						
4) Vinyl Chloride	1.374	62	10448	6.839	ug/l	93
16) Acetone	2.392	43	967	1.139	ug/l #	80
27) cis-1,2-Dichloroethene	4.483	96	6365	4.017	ug/l	89
44) Trichloroethene	7.123	130	49229	30.925	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112524\
Data File : VX043982.D
Acq On : 25 Nov 2024 13:31
Operator : JC/MD
Sample : P4846-01DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-MW-EVAL-03D-20241113DL

Quant Time: Nov 26 00:47:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

