

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112524\
 Data File : VX043984.D
 Acq On : 25 Nov 2024 14:17
 Operator : JC/MD
 Sample : P4886-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Storage-Blank-SOIL-REF-6

Quant Time: Nov 26 00:48:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	115325	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.757	114	219623	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	194128	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	78448	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	101733	51.432	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.860%
35) Dibromofluoromethane	5.379	113	74017	47.165	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.340%
50) Toluene-d8	8.647	98	268813	49.692	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.380%
62) 4-Bromofluorobenzene	11.079	95	88052	47.827	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	95.660%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112524\
Data File : VX043984.D
Acq On : 25 Nov 2024 14:17
Operator : JC/MD
Sample : P4886-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Storage-Blank-SOIL-REF-6

Quant Time: Nov 26 00:48:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

