

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112524\
 Data File : VX043986.D
 Acq On : 25 Nov 2024 15:03
 Operator : JC/MD
 Sample : P4886-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Storage-Blank-WATER-REF-4

Quant Time: Nov 26 00:48:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	111702	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.757	114	214805	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	187487	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	75443	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	96245	50.235	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	100.480%	
35) Dibromofluoromethane	5.379	113	71352	46.487	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	92.980%	
50) Toluene-d8	8.647	98	261482	49.421	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	98.840%	
62) 4-Bromofluorobenzene	11.079	95	85344	47.396	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	94.800%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112524\
Data File : VX043986.D
Acq On : 25 Nov 2024 15:03
Operator : JC/MD
Sample : P4886-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Storage-Blank-WATER-REF-4

Quant Time: Nov 26 00:48:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

