

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112524\
 Data File : VX043989.D
 Acq On : 25 Nov 2024 16:12
 Operator : JC/MD
 Sample : P4873-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-MW-19-20241114

Quant Time: Nov 26 00:50:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	122379	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.757	114	236964	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	204246	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	83659	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	108381	51.634	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 103.260%			
35) Dibromofluoromethane	5.379	113	79345	46.860	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 93.720%			
50) Toluene-d8	8.646	98	287176	49.202	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 98.400%			
62) 4-Bromofluorobenzene	11.079	95	91476	46.051	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 92.100%			
Target Compounds						
16) Acetone	2.379	43	1318	1.367	ug/l	89
44) Trichloroethene	7.122	130	9081	4.881	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112524\
Data File : VX043989.D
Acq On : 25 Nov 2024 16:12
Operator : JC/MD
Sample : P4873-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-MW-19-20241114

Quant Time: Nov 26 00:50:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

