

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112524\
 Data File : VX043990.D
 Acq On : 25 Nov 2024 16:35
 Operator : JC/MD
 Sample : P4873-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-MW-32-20241114

Quant Time: Nov 26 00:50:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	113361	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	214013	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	187289	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	77640	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	98009	50.407	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 100.820%			
35) Dibromofluoromethane	5.385	113	73131	47.822	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 95.640%			
50) Toluene-d8	8.647	98	260964	49.506	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 99.020%			
62) 4-Bromofluorobenzene	11.079	95	84479	47.090	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 94.180%			
Target Compounds						
16) Acetone	2.386	43	3293	3.687	ug/l #	89
27) cis-1,2-Dichloroethene	4.483	96	29183	17.506	ug/l	89
44) Trichloroethene	7.123	130	80066	47.648	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112524\
Data File : VX043990.D
Acq On : 25 Nov 2024 16:35
Operator : JC/MD
Sample : P4873-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-MW-32-20241114

Quant Time: Nov 26 00:50:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

