

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112524\
 Data File : VX043992.D
 Acq On : 25 Nov 2024 17:21
 Operator : JC/MD
 Sample : P4873-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-GEP-3-20241114

Quant Time: Nov 26 00:51:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	117514	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	213727	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	189569	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	78823	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	98843	49.040	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.080%
35) Dibromofluoromethane	5.385	113	74168	48.565	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.140%
50) Toluene-d8	8.647	98	263027	49.964	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.920%
62) 4-Bromofluorobenzene	11.079	95	86837	48.469	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	96.940%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	1513	1.634	ug/l	91
44) Trichloroethene	7.117	130	252075	150.213	ug/l	88
64) Tetrachloroethene	9.275	164	931	0.745	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112524\
Data File : VX043992.D
Acq On : 25 Nov 2024 17:21
Operator : JC/MD
Sample : P4873-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-GEP-3-20241114

Quant Time: Nov 26 00:51:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

