

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112524\
 Data File : VX044000.D
 Acq On : 25 Nov 2024 20:26
 Operator : JC/MD
 Sample : P4938-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-734

Quant Time: Nov 26 00:56:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	118735	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.757	114	225451	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	201840	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	88350	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	102246	50.207	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.420%
35) Dibromofluoromethane	5.385	113	76985	47.788	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.580%
50) Toluene-d8	8.647	98	274772	49.480	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.960%
62) 4-Bromofluorobenzene	11.079	95	92930	49.172	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	98.340%
Target Compounds						
16) Acetone	2.386	43	49431	52.848	ug/l	95
18) Methyl Acetate	2.703	43	3282	1.507	ug/l	95
20) Methylene Chloride	2.788	84	2329	1.467	ug/l	89
43) Isopropyl Acetate	6.342	43	119254	30.071	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112524\
Data File : VX044000.D
Acq On : 25 Nov 2024 20:26
Operator : JC/MD
Sample : P4938-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-734

Quant Time: Nov 26 00:56:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

