

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112524\
 Data File : VX044001.D
 Acq On : 25 Nov 2024 20:49
 Operator : JC/MD
 Sample : P4995-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 001

Quant Time: Nov 26 00:56:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	112737	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.757	114	221459	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	197663	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83986	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	98537	50.960	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.920%
35) Dibromofluoromethane	5.379	113	72939	46.093	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.180%
50) Toluene-d8	8.646	98	275478	50.502	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.000%
62) 4-Bromofluorobenzene	11.079	95	91730	49.412	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	98.820%
Target Compounds						
					Qvalue	
16) Acetone	2.385	43	2307	2.598	ug/l	90
18) Methyl Acetate	2.702	43	3479	1.682	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112524\
Data File : VX044001.D
Acq On : 25 Nov 2024 20:49
Operator : JC/MD
Sample : P4995-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
001

Quant Time: Nov 26 00:56:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration

