

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112525\
 Data File : VX048673.D
 Acq On : 25 Nov 2025 10:17
 Operator : JC/MD
 Sample : VX1125MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1125MBL01

Quant Time: Nov 26 00:57:01 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:08:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.531	168	113534	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.739	114	240284	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	270962	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	139976	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.934	65	97786	61.811	ug/l	-0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	=	123.620%#
35) Dibromofluoromethane	5.367	113	84870	46.667	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.340%
50) Toluene-d8	8.628	98	313982	55.357	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	110.720%
62) 4-Bromofluorobenzene	11.061	95	133106	62.971	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	125.940%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112525\
Data File : VX048673.D
Acq On : 25 Nov 2025 10:17
Operator : JC/MD
Sample : VX1125MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1125MBL01

Quant Time: Nov 26 00:57:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 05:08:11 2025
Response via : Initial Calibration

