

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112525\
 Data File : VX048679.D
 Acq On : 25 Nov 2025 13:00
 Operator : JC/MD
 Sample : Q3715-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW7

Quant Time: Nov 26 01:00:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:08:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	166997	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	305521	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	379381	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	188892	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	119607	51.400	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.800%
35) Dibromofluoromethane	5.373	113	108925	47.105	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.220%
50) Toluene-d8	8.635	98	397876	55.169	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	110.340%
62) 4-Bromofluorobenzene	11.061	95	161369	60.041	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	120.080%
Target Compounds						
						Qvalue
3) Chloromethane	1.301	50	930	0.518	ug/l	# 80
64) Tetrachloroethene	9.263	164	7016	2.546	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112525\
Data File : VX048679.D
Acq On : 25 Nov 2025 13:00
Operator : JC/MD
Sample : Q3715-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW7

Quant Time: Nov 26 01:00:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 05:08:11 2025
Response via : Initial Calibration

