

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX112618\
 Data File : VX006170.D
 Acq On : 26 Nov 2018 12:58
 Operator : JC/MD
 Sample : J6031-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Nov 27 02:58:21 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 21 08:15:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	326777	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	492038	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	453003	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	217785	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	206078	54.42	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.84%	
35) Dibromofluoromethane	5.50	113	165686	52.46	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.92%	
50) Toluene-d8	8.71	98	595103	49.40	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.80%	
62) 4-Bromofluorobenzene	11.14	95	198492	43.91	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.82%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX112618\
Data File : VX006170.D
Acq On : 26 Nov 2018 12:58
Operator : JC/MD
Sample : J6031-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Nov 27 02:58:21 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 21 08:15:54 2018
Response via : Initial Calibration

