

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX112618\
 Data File : VX006172.D
 Acq On : 26 Nov 2018 13:51
 Operator : JC/MD
 Sample : J6032-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Nov 27 03:04:13 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 21 08:15:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	330443	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	492643	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	451129	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	219236	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	206780	54.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.00%	
35) Dibromofluoromethane	5.50	113	166277	52.58	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.16%	
50) Toluene-d8	8.71	98	593955	49.24	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.48%	
62) 4-Bromofluorobenzene	11.14	95	199257	44.03	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.06%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX112618\
Data File : VX006172.D
Acq On : 26 Nov 2018 13:51
Operator : JC/MD
Sample : J6032-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Nov 27 03:04:13 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 21 08:15:54 2018
Response via : Initial Calibration

