

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX112618\
 Data File : VX006177.D
 Acq On : 26 Nov 2018 16:04
 Operator : JC/MD
 Sample : J6003-11 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 AOC1-03R

Quant Time: Nov 27 03:22:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 21 08:15:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	323026	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	483139	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	445040	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	220776	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	201646	53.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.74%	
35) Dibromofluoromethane	5.50	113	162593	52.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.86%	
50) Toluene-d8	8.71	98	587126	49.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.26%	
62) 4-Bromofluorobenzene	11.14	95	199605	44.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.94%	
Target Compounds						
16) Acetone	2.45	43	4378	2.64	ug/l	# 85
20) Methylene Chloride	2.86	84	3378	1.00	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX112618\
Data File : VX006177.D
Acq On : 26 Nov 2018 16:04
Operator : JC/MD
Sample : J6003-11 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
AOC1-03R

Quant Time: Nov 27 03:22:41 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 21 08:15:54 2018
Response via : Initial Calibration

