

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX112618\
 Data File : VX006181.D
 Acq On : 26 Nov 2018 17:51
 Operator : JC/MD
 Sample : PB115023TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB115023TB

Quant Time: Nov 27 03:39:36 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 21 08:15:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	313287	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	472974	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	460820	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	215801	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	199290	54.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.80%	
35) Dibromofluoromethane	5.50	113	160097	52.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.46%	
50) Toluene-d8	8.71	98	607709	52.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.96%	
62) 4-Bromofluorobenzene	11.14	95	204177	46.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.98%	
Target Compounds						
16) Acetone	2.45	43	11929	7.41	ug/l	99
20) Methylene Chloride	2.86	84	11186	4.27	ug/l	93
43) Isopropyl Acetate	6.46	43	20332	2.54	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX112618\
Data File : VX006181.D
Acq On : 26 Nov 2018 17:51
Operator : JC/MD
Sample : PB115023TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB115023TB

Quant Time: Nov 27 03:39:36 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 21 08:15:54 2018
Response via : Initial Calibration

