

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX112618\
 Data File : VX006184.D
 Acq On : 26 Nov 2018 19:11
 Operator : JC/MD
 Sample : PB115023ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB115023ZHE#03

Quant Time: Nov 27 03:47:12 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 21 08:15:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	309980	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	466687	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	461244	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	210688	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	197173	54.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.78%	
35) Dibromofluoromethane	5.51	113	161040	53.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.52%	
50) Toluene-d8	8.71	98	604727	52.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.84%	
62) 4-Bromofluorobenzene	11.14	95	197545	46.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.16%	
Target Compounds						
16) Acetone	2.45	43	13623	8.55	ug/l	99
18) Methyl Acetate	2.78	43	6365	1.56	ug/l #	87
20) Methylene Chloride	2.86	84	10955	4.22	ug/l	97
43) Isopropyl Acetate	6.46	43	19863	2.52	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX112618\
Data File : VX006184.D
Acq On : 26 Nov 2018 19:11
Operator : JC/MD
Sample : PB115023ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB115023ZHE#03

Quant Time: Nov 27 03:47:12 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 21 08:15:54 2018
Response via : Initial Calibration

