

Data File : VX006186.D
Acq On : 26 Nov 2018 20:05
Operator : JC/MD
Sample : J5999-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SSSI-1114-S-TCLP-3

Quant Time: Nov 27 03:52:14 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 21 08:15:54 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	305510	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	455269	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	453672	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	208783	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	194717	55.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.00%	
35) Dibromofluoromethane	5.50	113	155408	53.18	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.36%	
50) Toluene-d8	8.71	98	595038	53.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.76%	
62) 4-Bromofluorobenzene	11.14	95	197031	47.11	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.22%	

Target Compounds				Qvalue		
16) Acetone	2.45	43	12955	8.247	ug/l	96
18) Methyl Acetate	2.78	43	7106	1.764	ug/l	98
20) Methylene Chloride	2.85	84	10980	4.300	ug/l	97
43) Isopropyl Acetate	6.46	43	18435	2.396	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX006186.D
Acq On : 26 Nov 2018 20:05
Operator : JC/MD
Sample : J5999-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SSSI-1114-S-TCLP-3

Quant Time: Nov 27 03:52:14 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 21 08:15:54 2018
Response via : Initial Calibration

