

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX112618\
 Data File : VX006194.D
 Acq On : 27 Nov 2018 00:31
 Operator : JC/MD
 Sample : PB115023ZHE#07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB115023ZHE#07

Quant Time: Nov 27 04:32:18 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 21 08:15:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	294357	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	437347	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	435100	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	201009	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	188422	55.24	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.48%	
35) Dibromofluoromethane	5.51	113	150050	53.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.90%	
50) Toluene-d8	8.71	98	574739	53.67	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.34%	
62) 4-Bromofluorobenzene	11.14	95	187543	46.68	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.36%	

Target Compounds						Qvalue
16) Acetone	2.45	43	12927	8.54	ug/l	100
20) Methylene Chloride	2.85	84	12234	5.03	ug/l	93
43) Isopropyl Acetate	6.46	43	18697	2.53	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX112618\
Data File : VX006194.D
Acq On : 27 Nov 2018 00:31
Operator : JC/MD
Sample : PB115023ZHE#07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB115023ZHE#07

Quant Time: Nov 27 04:32:18 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 21 08:15:54 2018
Response via : Initial Calibration

