

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX112618\
 Data File : VX006195.D
 Acq On : 27 Nov 2018 00:57
 Operator : JC/MD
 Sample : PB115023ZHE#08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB115023ZHE#08

Quant Time: Nov 27 04:35:07 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 21 08:15:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	264537	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	403647	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	403713	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	184339	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	176839	57.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.38%	
35) Dibromofluoromethane	5.50	113	139774	53.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.88%	
50) Toluene-d8	8.71	98	529842	53.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.22%	
62) 4-Bromofluorobenzene	11.14	95	172631	46.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.12%	
Target Compounds						
16) Acetone	2.44	43	9204	6.77	ug/l	93
20) Methylene Chloride	2.86	84	4038	1.62	ug/l	96
43) Isopropyl Acetate	6.46	43	17392	2.55	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX112618\
Data File : VX006195.D
Acq On : 27 Nov 2018 00:57
Operator : JC/MD
Sample : PB115023ZHE#08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB115023ZHE#08

Quant Time: Nov 27 04:35:07 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 21 08:15:54 2018
Response via : Initial Calibration

