

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX112618\
 Data File : VX006197.D
 Acq On : 27 Nov 2018 01:51
 Operator : JC/MD
 Sample : J5770-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OR-02-112118-B

Quant Time: Nov 27 04:40:15 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 21 08:15:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	261131	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	392390	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	395224	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	185477	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	172274	56.93	ug/l	0.00
Spiked Amount	50.000		Recovery	=	113.86%	
35) Dibromofluoromethane	5.50	113	137495	54.59	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.18%	
50) Toluene-d8	8.71	98	522660	54.40	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.80%	
62) 4-Bromofluorobenzene	11.14	95	169594	47.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.10%	

Target Compounds						Qvalue
16) Acetone	2.45	43	43960	32.74	ug/l	97
18) Methyl Acetate	2.78	43	9432	2.74	ug/l	# 91
20) Methylene Chloride	2.85	84	13291	6.24	ug/l	93
43) Isopropyl Acetate	6.46	43	14841	2.24	ug/l	96
95) Naphthalene	13.83	128	40575	3.31	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX112618\
Data File : VX006197.D
Acq On : 27 Nov 2018 01:51
Operator : JC/MD
Sample : J5770-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OR-02-112118-B

Quant Time: Nov 27 04:40:15 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 21 08:15:54 2018
Response via : Initial Calibration

