

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112618\
 Data File : VX006203.D
 Acq On : 27 Nov 2018 04:31
 Operator : JC/MD
 Sample : PB115023ZHE#06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB115023ZHE#06

Quant Time: Nov 27 05:04:37 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 21 08:15:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	265663	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	402691	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	393779	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	185275	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	174794	56.78	ug/l	0.00
Spiked Amount			Recovery	=	113.56%	
35) Dibromofluoromethane	5.50	113	137916	53.35	ug/l	0.00
Spiked Amount			Recovery	=	106.70%	
50) Toluene-d8	8.71	98	514632	52.20	ug/l	0.00
Spiked Amount			Recovery	=	104.40%	
62) 4-Bromofluorobenzene	11.14	95	169422	45.80	ug/l	0.00
Spiked Amount			Recovery	=	91.60%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	17644	12.916	ug/l	93
18) Methyl Acetate	2.78	43	3715	1.061	ug/l #	74
20) Methylene Chloride	2.86	84	115248	55.781	ug/l	95
43) Isopropyl Acetate	6.46	43	7042	1.035	ug/l #	96
95) Naphthalene	13.83	128	19218	1.567	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112618\
Data File : VX006203.D
Acq On : 27 Nov 2018 04:31
Operator : JC/MD
Sample : PB115023ZHE#06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB115023ZHE#06

Quant Time: Nov 27 05:04:37 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 21 08:15:54 2018
Response via : Initial Calibration

