

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX112618\
 Data File : VX006205.D
 Acq On : 27 Nov 2018 05:24
 Operator : JC/MD
 Sample : J6050-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DTW-02

Quant Time: Nov 27 08:14:51 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 21 08:15:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	269149	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	399220	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	375192	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	188907	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	175115	56.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.30%	
35) Dibromofluoromethane	5.50	113	138903	54.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.40%	
50) Toluene-d8	8.71	98	491044	50.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.48%	
62) 4-Bromofluorobenzene	11.14	95	166984	45.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.06%	
Target Compounds						
16) Acetone	2.45	43	22262	16.09	ug/l	99
20) Methylene Chloride	2.86	84	3574	1.37	ug/l	96
95) Naphthalene	13.83	128	26203	2.10	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX112618\
Data File : VX006205.D
Acq On : 27 Nov 2018 05:24
Operator : JC/MD
Sample : J6050-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DTW-02

Quant Time: Nov 27 08:14:51 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 21 08:15:54 2018
Response via : Initial Calibration

