

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX112618\
 Data File : VX006206.D
 Acq On : 27 Nov 2018 05:51
 Operator : JC/MD
 Sample : J6059-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OUTFALL001-181120

Quant Time: Nov 27 09:42:00 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 21 08:15:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	277359	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	415489	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	381242	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	186399	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	178082	55.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.82%	
35) Dibromofluoromethane	5.50	113	142236	53.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.66%	
50) Toluene-d8	8.71	98	505989	49.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.48%	
62) 4-Bromofluorobenzene	11.13	95	168877	44.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.50%	
Target Compounds						
16) Acetone	2.44	43	13079	9.17	ug/l	96
27) cis-1,2-Dichloroethene	4.60	96	11721	4.00	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX112618\
Data File : VX006206.D
Acq On : 27 Nov 2018 05:51
Operator : JC/MD
Sample : J6059-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OUTFALL001-181120

Quant Time: Nov 27 09:42:00 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 21 08:15:54 2018
Response via : Initial Calibration

