

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX112618\
 Data File : VX006206.D
 Acq On : 27 Nov 2018 05:51
 Operator : JC/MD
 Sample : J6059-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OUTFALL001-181120

Quant Time: Nov 27 09:42:00 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 21 08:15:54 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	277359	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	415489	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	381242	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	186399	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	178082	55.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.82%	
35) Dibromofluoromethane	5.50	113	142236	53.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.66%	
50) Toluene-d8	8.71	98	505989	49.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.48%	
62) 4-Bromofluorobenzene	11.13	95	168877	44.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.50%	
Target Compounds						
16) Acetone	2.44	43	13079	9.17	ug/l	96
27) cis-1,2-Dichloroethene	4.60	96	11721	4.00	ug/l	90


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX112618\
Data File : VX006206.D
Acq On : 27 Nov 2018 05:51
Operator : JC/MD
Sample : J6059-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OUTFALL001-181120

Quant Time: Nov 27 09:42:00 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112018W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 21 08:15:54 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 5.67 min Scan# 740

Delta R.T. 0.00 min

Lab File: VX006206.D

Acq: 27 Nov 2018 05:51

Instrument :

MSVOA_X

ClientSampleId :


OUTFALL001-181120

Tgt Ion:168 Resp: 277359

Ion Ratio Lower Upper




168 100

99 51.1 39.2 58.8



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#16

Acetone

Concen: 9.17 ug/l

RT: 2.44 min Scan# 211

Delta R.T. -0.01 min

Lab File: VX006206.D


Acq: 27 Nov 2018 05:51

Tgt Ion: 43 Resp: 13079

Ion Ratio Lower Upper



43 100


58 27.7 23.8 35.8



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0





#27

cis-1,2-Dichloroethene

Concen: 4.00 ug/l

RT: 4.60 min Scan# 566

Delta R.T. 0.01 min

Lab File: VX006206.D

Acq: 27 Nov 2018 05:51

Instrument :

MSVOA_X

ClientSampleId :

OUTFALL001-181120


Tgt Ion: 96 Resp: 11721

Ion Ratio Lower Upper

96 100

61 130.2 0.0 291.6




98 59.9 0.0 128.2



Abundance Ion 95.90 (95.60 to 96.60): VX0

Ion 60.90 (60.60 to 61.60): VX0

Ion 97.90 (97.60 to 98.60): VX0



#33

1,2-Dichloroethane-d4

Concen: 55.41 ug/l

RT: 6.07 min Scan# 806

Delta R.T. 0.00 min

Lab File: VX006206.D


Acq: 27 Nov 2018 05:51

Tgt Ion: 65 Resp: 178082

Ion Ratio Lower Upper



65 100


67 52.2 0.0 105.8



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0





#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 6.86 min Scan# 936

Delta R.T. 0.00 min

Lab File: VX006206.D

Acq: 27 Nov 2018 05:51

Instrument :

MSVOA_X

ClientSampleId :

OUTFALL001-181120


Tgt Ion:114 Resp: 415489

Ion Ratio Lower Upper

114 100

63 20.7 0.0 39.2




88 14.8 0.0 28.6



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 53.33 ug/l

RT: 5.50 min Scan# 713

Delta R.T. 0.00 min

Lab File: VX006206.D

Acq: 27 Nov 2018 05:51


Tgt Ion:113 Resp: 142236

Ion Ratio Lower Upper

113 100

111 102.5 81.5 122.3



192 22.5 18.6 28.0




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 49.74 ug/l

RT: 8.71 min Scan# 1240

Delta R.T. 0.00 min

Lab File: VX006206.D

Acq: 27 Nov 2018 05:51

Instrument :

MSVOA_X

ClientSampleId :


OUTFALL001-181120

Tgt Ion: 98 Resp: 505989

Ion Ratio Lower Upper




98 100

100 63.2 51.0 76.4



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): V



#62

4-Bromofluorobenzene

Concen: 44.25 ug/l

RT: 11.13 min Scan# 1637

Delta R.T. -0.01 min

Lab File: VX006206.D

Acq: 27 Nov 2018 05:51


Tgt Ion: 95 Resp: 168877

Ion Ratio Lower Upper

95 100

174 94.4 0.0 184.4



176 89.8 0.0 179.0




Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.12 min Scan# 1470
Delta R.T. 0.00 min
Lab File: VX006206.D
Acq: 27 Nov 2018 05:51

Instrument :
MSVOA_X
ClientSampleId :
OUTFALL001-181120

Tgt Ion:117 Resp: 381242
Ion Ratio Lower Upper
117 100
82 54.9 43.0 64.4
119 32.5 26.1 39.1




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.08 min Scan# 1792
Delta R.T. 0.00 min
Lab File: VX006206.D
Acq: 27 Nov 2018 05:51

Tgt Ion:152 Resp: 186399
Ion Ratio Lower Upper
152 100
115 55.2 39.0 116.9
150 156.5 0.0 346.6



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

