

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112619\
 Data File : VX013604.D
 Acq On : 26 Nov 2019 17:17
 Operator : JC/SP
 Sample : VX1126WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1126WBL01

Quant Time: Nov 26 17:57:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	675215	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1083438	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	955305	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	422982	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	392480	50.08	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	100.16%	
35) Dibromofluoromethane	5.48	113	325253	48.56	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	97.12%	
50) Toluene-d8	8.71	98	1283469	49.18	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	98.36%	
62) 4-Bromofluorobenzene	11.13	95	429850	45.68	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	91.36%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112619\
Data File : VX013604.D
Acq On : 26 Nov 2019 17:17
Operator : JC/SP
Sample : VX1126WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1126WBL01

Quant Time: Nov 26 17:57:48 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

