

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112619\
 Data File : VX013605.D
 Acq On : 26 Nov 2019 17:47
 Operator : JC/SP
 Sample : K5979-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 391120791-TRIP-BLANK

Quant Time: Nov 26 18:09:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	712040	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1156565	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1029466	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	472326	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	414352	50.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.26%	
35) Dibromofluoromethane	5.48	113	345594	48.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.66%	
50) Toluene-d8	8.71	98	1381974	49.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.22%	
62) 4-Bromofluorobenzene	11.13	95	461837	45.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.94%	
Target Compounds						
16) Acetone	2.43	43	598932	146.775	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112619\
Data File : VX013605.D
Acq On : 26 Nov 2019 17:47
Operator : JC/SP
Sample : K5979-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
391120791-TRIP-BLANK

Quant Time: Nov 26 18:09:54 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

