

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112619\
 Data File : VX013609.D
 Acq On : 26 Nov 2019 19:20
 Operator : JC/SP
 Sample : K6002-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-15

Quant Time: Nov 27 04:06:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	664619	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1065207	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	938876	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	414956	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	383386	49.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.40%	
35) Dibromofluoromethane	5.49	113	317546	48.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.44%	
50) Toluene-d8	8.71	98	1258519	49.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.10%	
62) 4-Bromofluorobenzene	11.14	95	417362	45.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.22%	
Target Compounds						
30) Chloroform	5.20	83	8915	0.728	ug/l	90
44) Trichloroethene	7.22	130	4082	0.505	ug/l	95
64) Tetrachloroethene	9.33	164	28657	4.093	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112619\
Data File : VX013609.D
Acq On : 26 Nov 2019 19:20
Operator : JC/SP
Sample : K6002-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-15

Quant Time: Nov 27 04:06:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

