

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112619\
 Data File : VX013622.D
 Acq On : 27 Nov 2019 01:09
 Operator : JC/SP
 Sample : K5962-03DL 100X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 15BR-20191120DL

Quant Time: Nov 27 05:54:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	688518	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1091147	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	980699	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	439733	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	394033	49.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.60%	
35) Dibromofluoromethane	5.49	113	326001	48.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.66%	
50) Toluene-d8	8.71	98	1309808	49.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.68%	
62) 4-Bromofluorobenzene	11.14	95	438569	46.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.56%	
Target Compounds						
4) Vinyl Chloride	1.40	62	14215	1.469	ug/l	97
27) cis-1,2-Dichloroethene	4.59	96	201906	24.630	ug/l	90
44) Trichloroethene	7.21	130	222981	26.933	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112619\
Data File : VX013622.D
Acq On : 27 Nov 2019 01:09
Operator : JC/SP
Sample : K5962-03DL 100X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
15BR-20191120DL

Quant Time: Nov 27 05:54:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

