

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112619\
 Data File : VX013624.D
 Acq On : 27 Nov 2019 01:55
 Operator : JC/SP
 Sample : K5962-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 48BR-20191120

Quant Time: Nov 27 06:04:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	638910	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1020169	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	909893	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	406277	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	371715	50.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.24%	
35) Dibromofluoromethane	5.48	113	306358	48.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.16%	
50) Toluene-d8	8.71	98	1209066	49.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.40%	
62) 4-Bromofluorobenzene	11.14	95	405294	45.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.48%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.40	62	6953	0.774	ug/l #	87
9) 1,1,2-Trichlorotrifluoroet	2.38	101	11930	1.958	ug/l	91
12) 1,1-Dichloroethene	2.37	96	33422	5.497	ug/l	91
16) Acetone	2.44	43	7805	2.132	ug/l #	76
21) trans-1,2-Dichloroethene	3.16	96	3631	0.548	ug/l #	89
27) cis-1,2-Dichloroethene	4.59	96	148777	19.558	ug/l	88
44) Trichloroethene	7.21	130	4929165	636.787	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112619\
Data File : VX013624.D
Acq On : 27 Nov 2019 01:55
Operator : JC/SP
Sample : K5962-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
48BR-20191120

Quant Time: Nov 27 06:04:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

