

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112619\
 Data File : VX013627.D
 Acq On : 27 Nov 2019 03:05
 Operator : JC/SP
 Sample : K5962-12 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 91BR-20191120

Quant Time: Nov 27 06:20:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	656361	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1045628	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	929145	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	416988	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	381144	50.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.06%	
35) Dibromofluoromethane	5.48	113	317298	49.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.18%	
50) Toluene-d8	8.71	98	1244243	49.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.80%	
62) 4-Bromofluorobenzene	11.14	95	415048	45.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.40%	
Target Compounds						
4) Vinyl Chloride	1.40	62	191436	20.749	ug/l	98
11) Tert butyl alcohol	3.01	59	21650	11.512	ug/l #	74
27) cis-1,2-Dichloroethene	4.59	96	891504	114.082	ug/l	89
44) Trichloroethene	7.21	130	59098	7.449	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112619\
Data File : VX013627.D
Acq On : 27 Nov 2019 03:05
Operator : JC/SP
Sample : K5962-12 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
91BR-20191120

Quant Time: Nov 27 06:20:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

