

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112619\
 Data File : VX013629.D
 Acq On : 27 Nov 2019 03:52
 Operator : JC/SP
 Sample : K5968-14
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ROOF-DRAIN-20191121

Quant Time: Nov 27 06:30:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	642536	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1026944	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	925631	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	409545	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	379004	50.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.64%	
35) Dibromofluoromethane	5.49	113	307006	48.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.72%	
50) Toluene-d8	8.71	98	1225403	49.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.08%	
62) 4-Bromofluorobenzene	11.14	95	408917	45.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.68%	
Target Compounds						
16) Acetone	2.44	43	14135	3.839	ug/l	92
44) Trichloroethene	7.22	130	5132	0.659	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112619\
Data File : VX013629.D
Acq On : 27 Nov 2019 03:52
Operator : JC/SP
Sample : K5968-14
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ROOF-DRAIN-20191121

Quant Time: Nov 27 06:30:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

