

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112619\
 Data File : VX013631.D
 Acq On : 27 Nov 2019 04:38
 Operator : JC/SP
 Sample : K5968-19
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-01-20191121

Quant Time: Nov 27 06:39:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	657410	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1052968	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	936665	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	420963	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	390467	51.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.34%	
35) Dibromofluoromethane	5.49	113	317114	48.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.44%	
50) Toluene-d8	8.71	98	1266309	49.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.86%	
62) 4-Bromofluorobenzene	11.14	95	418093	45.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.44%	
Target Compounds						
44) Trichloroethene	7.22	130	3047	0.381	ug/l	Qvalue 69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112619\
Data File : VX013631.D
Acq On : 27 Nov 2019 04:38
Operator : JC/SP
Sample : K5968-19
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-01-20191121

Quant Time: Nov 27 06:39:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

