

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX112619\
 Data File : VX013636.D
 Acq On : 27 Nov 2019 06:35
 Operator : JC/SP
 Sample : K6002-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-10D

Quant Time: Nov 27 07:31:59 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	674357	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1072549	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	955660	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	435740	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	385957	49.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.62%	
35) Dibromofluoromethane	5.49	113	318122	47.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.96%	
50) Toluene-d8	8.71	98	1275347	49.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.74%	
62) 4-Bromofluorobenzene	11.14	95	426960	45.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.66%	
Target Compounds						
					Qvalue	
30) Chloroform	5.20	83	6384	0.513	ug/l #	74
40) Benzene	6.13	78	8163	0.275	ug/l #	81
64) Tetrachloroethene	9.33	164	95958	13.465	ug/l	97
67) Ethyl Benzene	10.25	91	8334	0.244	ug/l	97
68) m/p-Xylenes	10.36	106	13845	1.061	ug/l	91
69) o-Xylene	10.70	106	3608	0.283	ug/l	83
84) 1,2,4-Trimethylbenzene	11.81	105	9368	0.368	ug/l	95
95) Naphthalene	13.83	128	71659	2.428	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX112619\
Data File : VX013636.D
Acq On : 27 Nov 2019 06:35
Operator : JC/SP
Sample : K6002-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-10D

Quant Time: Nov 27 07:31:59 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

