

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX112619\
 Data File : VX013613.D
 Acq On : 26 Nov 2019 20:53
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 27 04:30:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	97	0.00
2 T	Dichlorodifluoromethane	0.557	0.564	-1.3	97	0.00
3 P	Chloromethane	0.624	0.611	2.1	96	0.00
4 C	Vinyl Chloride	0.703	0.717	-2.0#	98	0.00
5 T	Bromomethane	0.472	0.438	7.2	98	0.00
6 T	Chloroethane	0.400	0.382	4.5	98	0.00
7 T	Trichlorofluoromethane	0.780	0.754	3.3	99	0.00
8 T	Diethyl Ether	0.353	0.346	2.0	99	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.477	0.458	4.0	99	0.00
10 T	Methyl Iodide	0.577	0.605	-4.9	97	0.00
11 T	Tert butyl alcohol	0.143	0.128	10.5	94	0.00
12 CM	1,1-Dichloroethene	0.476	0.472	0.8#	99	0.00
13 T	Acrolein	0.088	0.088	0.0	90	0.00
14 T	Allyl chloride	0.864	0.841	2.7	96	0.00
15 T	Acrylonitrile	0.296	0.288	2.7	97	0.00
16 T	Acetone	0.287	0.257	10.5	89	0.00
17 T	Carbon Disulfide	1.356	1.297	4.4	98	0.00
18 T	Methyl Acetate	0.798	0.815	-2.1	98	0.00
19 T	Methyl tert-butyl Ether	1.576	1.585	-0.6	100	0.00
20 T	Methylene Chloride	0.551	0.525	4.7	99	0.00
21 T	trans-1,2-Dichloroethene	0.518	0.507	2.1	99	0.00
22 T	Diisopropyl ether	1.663	1.656	0.4	98	0.00
23 T	Vinyl Acetate	1.383	1.398	-1.1	98	0.00
24 P	1,1-Dichloroethane	0.909	0.907	0.2	99	0.00
25 T	2-Butanone	0.426	0.411	3.5	93	0.00
26 T	2,2-Dichloropropane	0.752	0.707	6.0	94	0.00
27 T	cis-1,2-Dichloroethene	0.595	0.578	2.9	98	0.00
28 T	Bromochloromethane	0.320	0.369	-15.3	99	0.00
29 T	Tetrahydrofuran	0.263	0.263	0.0	96	0.00
30 C	Chloroform	0.922	0.883	4.2#	98	0.00
31 T	Cyclohexane	0.857	0.831	3.0	98	0.00
32 T	1,1,1-Trichloroethane	0.770	0.762	1.0	99	0.00
33 S	1,2-Dichloroethane-d4	0.580	0.563	2.9	100	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
35 S	Dibromofluoromethane	0.309	0.305	1.3	102	0.00
36 T	1,1-Dichloropropene	0.454	0.441	2.9	98	0.00
37 T	Ethyl Acetate	0.509	0.507	0.4	96	0.00
38 T	Carbon Tetrachloride	0.419	0.423	-1.0	100	0.00
39 T	Methylcyclohexane	0.561	0.546	2.7	99	0.00
40 TM	Benzene	1.382	1.360	1.6	100	0.00
41 T	Methacrylonitrile	0.279	0.284	-1.8	97	0.00
42 TM	1,2-Dichloroethane	0.471	0.463	1.7	100	0.00
43 T	Isopropyl Acetate	0.838	0.839	-0.1	99	0.00
44 TM	Trichloroethene	0.379	0.365	3.7	98	0.00
45 C	1,2-Dichloropropane	0.346	0.345	0.3#	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX112619\
 Data File : VX013613.D
 Acq On : 26 Nov 2019 20:53
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 27 04:30:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 26 15:53:00 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.235	0.227	3.4	99	0.00
47 T	Bromodichloromethane	0.444	0.449	-1.1	100	0.00
48 T	Methyl methacrylate	0.407	0.416	-2.2	98	0.00
49 T	1,4-Dioxane	0.009	0.008	11.1	92	0.00
50 S	Toluene-d8	1.204	1.162	3.5	100	0.00
51 T	4-Methyl-2-Pentanone	0.527	0.519	1.5	97	0.00
52 CM	Toluene	0.859	0.852	0.8#	100	0.00
53 T	t-1,3-Dichloropropene	0.505	0.504	0.2	97	0.00
54 T	cis-1,3-Dichloropropene	0.552	0.560	-1.4	98	0.00
55 T	1,1,2-Trichloroethane	0.346	0.342	1.2	99	0.00
56 T	Ethyl methacrylate	0.556	0.565	-1.6	100	0.00
57 T	1,3-Dichloropropane	0.597	0.574	3.9	98	0.00
58 T	2-Chloroethyl Vinyl ether	0.218	0.214	1.8	98	0.00
59 T	2-Hexanone	0.411	0.400	2.7	95	0.00
60 T	Dibromochloromethane	0.359	0.366	-1.9	99	0.00
61 T	1,2-Dibromoethane	0.370	0.356	3.8	98	0.00
62 S	4-Bromofluorobenzene	0.434	0.421	3.0	101	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	96	0.00
64 T	Tetrachloroethene	0.373	0.369	1.1	103	0.00
65 PM	Chlorobenzene	1.038	1.014	2.3	99	0.00
66 T	1,1,1,2-Tetrachloroethane	0.373	0.377	-1.1	99	0.00
67 C	Ethyl Benzene	1.784	1.814	-1.7#	99	0.00
68 T	m/p-Xylenes	0.683	0.712	-4.2	103	0.00
69 T	o-Xylene	0.666	0.676	-1.5	99	0.00
70 T	Styrene	1.120	1.140	-1.8	98	0.00
71 P	Bromoform	0.308	0.325	-5.5	100	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
73 T	Isopropylbenzene	3.507	3.521	-0.4	99	0.00
74 T	N-amyl acetate	1.607	1.651	-2.7	97	0.00
75 P	1,1,2,2-Tetrachloroethane	1.216	1.197	1.6	97	0.00
76 T	1,2,3-Trichloropropane	1.125	0.968	14.0	83	0.00
77 T	Bromobenzene	0.948	0.925	2.4	99	0.00
78 T	n-propylbenzene	3.915	3.938	-0.6	98	0.00
79 T	2-Chlorotoluene	2.349	2.335	0.6	99	0.00
80 T	1,3,5-Trimethylbenzene	2.875	2.971	-3.3	101	0.00
81 T	trans-1,4-Dichloro-2-butene	0.411	0.412	-0.2	97	0.00
82 T	4-Chlorotoluene	2.720	2.685	1.3	99	0.00
83 T	tert-Butylbenzene	2.921	2.965	-1.5	100	0.00
84 T	1,2,4-Trimethylbenzene	2.917	3.057	-4.8	103	0.00
85 T	sec-Butylbenzene	3.337	3.358	-0.6	98	0.00
86 T	p-Isopropyltoluene	3.104	3.150	-1.5	99	0.00
87 T	1,3-Dichlorobenzene	1.652	1.616	2.2	98	0.00
88 T	1,4-Dichlorobenzene	1.682	1.625	3.4	100	0.00
89 T	n-Butylbenzene	2.648	2.723	-2.8	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX112619\
Data File : VX013613.D
Acq On : 26 Nov 2019 20:53
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Nov 27 04:30:39 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 26 15:53:00 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.561	0.561	0.0	100	0.00
91 T	1,2-Dichlorobenzene	1.632	1.621	0.7	99	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.265	0.268	-1.1	97	0.00
93 T	1,2,4-Trichlorobenzene	1.137	1.144	-0.6	100	0.00
94 T	Hexachlorobutadiene	0.546	0.511	6.4	100	0.00
95 T	Naphthalene	3.387	3.640	-7.5	102	0.00
96 T	1,2,3-Trichlorobenzene	1.126	1.135	-0.8	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6