


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX112619\
Data File : VX013595.D
Acq On : 26 Nov 2019 09:19
Operator : JC/SP
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112619W.M
Title : SW846 8260
Last Update : Tue Nov 26 15:53:00 2019



AutoFind: Scans 1679, 1680, 1681; Background Corrected with Scan 1673

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.4	35221	PASS
75	95	30	60	48.7	88493	PASS
95	95	100	100	100.0	181644	PASS
96	95	5	9	6.7	12172	PASS
173	174	0.00	2	0.9	1417	PASS
174	95	50	100	84.3	153045	PASS
175	174	5	9	7.5	11421	PASS
176	174	95	101	97.3	148917	PASS
177	176	5	9	6.3	9409	PASS