

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112621\
 Data File : VX025350.D
 Acq On : 26 Nov 2021 12:52
 Operator : JC/MD
 Sample : M4839-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ESQN5

Quant Time: Nov 26 23:57:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML112221WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Nov 26 23:56:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	146312	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	131292	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	61162	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.367	65	42554	42.617	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	85.240%
7) Chloroethane-d5	1.666	69	38463	46.672	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.340%
11) 1,1-Dichloroethene-d2	2.306	63	55982	35.163	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.320%
21) 2-Butanone-d5	4.458	46	82563	111.078	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	111.080%
24) Chloroform-d	5.062	84	86762	49.717	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.440%
26) 1,2-Dichloroethane-d4	5.958	65	55709	51.215	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.420%
32) Benzene-d6	5.976	84	176012	50.583	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.160%
36) 1,2-Dichloropropane-d6	7.312	67	54185	50.295	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.600%
41) Toluene-d8	8.653	98	158239	48.179	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.360%
43) trans-1,3-Dichloroprop...	8.951	79	26007	48.015	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.020%
47) 2-Hexanone-d5	9.384	63	60445	107.784	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	107.780%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	79309	52.088	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	104.180%
66) 1,2-Dichlorobenzene-d4	12.323	152	62089	51.758	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.520%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112621\
Data File : VX025350.D
Acq On : 26 Nov 2021 12:52
Operator : JC/MD
Sample : M4839-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ESQN5

Quant Time: Nov 26 23:57:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM112221WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Nov 26 23:56:11 2021
Response via : Initial Calibration

